

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
 Data File : VY014867.D
 Acq On : 26 Jul 2023 22:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 27 01:15:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	59	0.00
2 T	Dichlorodifluoromethane	0.407	0.308	24.3	47#	0.00
3 P	Chloromethane	0.458	0.457	0.2	60	0.00
4 C	Vinyl Chloride	0.512	0.531	-3.7#	61	0.00
5 T	Bromomethane	0.338	0.441	-30.5#	78	0.00
6 T	Chloroethane	0.320	0.373	-16.6	68	0.00
7 T	Trichlorofluoromethane	0.860	0.788	8.4	53	0.00
8 T	Diethyl Ether	0.320	0.312	2.5	57	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.470	10.8	51	0.00
10 T	Methyl Iodide	0.493	0.436	11.6	48#	0.00
11 T	Tert butyl alcohol	0.125	0.107	14.4	57	0.00
12 CM	1,1-Dichloroethene	0.492	0.431	12.4#	50	0.00
13 T	Acrolein	0.060	0.022	63.3#	24#	0.00
14 T	Allyl chloride	0.874	0.720	17.6	48#	0.00
15 T	Acrylonitrile	0.176	0.189	-7.4	62	0.00
16 T	Acetone	0.181	0.153	15.5	46#	0.00
17 T	Carbon Disulfide	1.410	1.088	22.8	46#	0.00
18 T	Methyl Acetate	0.693	0.784	-13.1	66	0.00
19 T	Methyl tert-butyl Ether	1.599	1.609	-0.6	58	0.00
20 T	Methylene Chloride	0.656	0.603	8.1	59	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.512	8.4	53	0.00
22 T	Diisopropyl ether	1.777	1.682	5.3	54	0.00
23 T	Vinyl Acetate	1.064	0.998	6.2	54	0.00
24 P	1,1-Dichloroethane	1.017	0.958	5.8	54	0.00
25 T	2-Butanone	0.264	0.263	0.4	57	0.00
26 T	2,2-Dichloropropane	0.969	0.772	20.3	46#	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.635	2.3	56	0.00
28 T	Bromochloromethane	0.438	0.403	8.0	57	0.00
29 T	Tetrahydrofuran	0.163	0.175	-7.4	64	0.00
30 C	Chloroform	1.049	1.042	0.7#	58	0.00
31 T	Cyclohexane	0.946	0.734	22.4	47#	0.00
32 T	1,1,1-Trichloroethane	0.944	0.891	5.6	55	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.545	4.4	58	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	62	0.00
35 S	Dibromofluoromethane	0.318	0.290	8.8	55	0.00
36 T	1,1-Dichloropropene	0.476	0.417	12.4	52	0.00
37 T	Ethyl Acetate	0.328	0.330	-0.6	63	0.00
38 T	Carbon Tetrachloride	0.503	0.457	9.1	54	0.00
39 T	Methylcyclohexane	0.620	0.509	17.9	49#	0.00
40 TM	Benzene	1.354	1.252	7.5	55	0.00
41 T	Methacrylonitrile	0.169	0.174	-3.0	64	0.00
42 TM	1,2-Dichloroethane	0.429	0.427	0.5	59	0.00
43 T	Isopropyl Acetate	0.624	0.608	2.6	58	0.00
44 TM	Trichloroethene	0.379	0.355	6.3	56	0.00
45 C	1,2-Dichloropropane	0.353	0.327	7.4#	55	0.00
46 T	Dibromomethane	0.214	0.216	-0.9	60	0.00
47 T	Bromodichloromethane	0.505	0.496	1.8	58	0.00
48 T	Methyl methacrylate	0.281	0.276	1.8	59	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
 Data File : VY014867.D
 Acq On : 26 Jul 2023 22:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 27 01:15:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	63	0.00
50 S	Toluene-d8	1.173	1.030	12.2	54	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.349	-6.4	64	0.00
52 CM	Toluene	0.866	0.811	6.4#	56	0.00
53 T	t-1,3-Dichloropropene	0.553	0.502	9.2	54	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.543	9.7	53	0.00
55 T	1,1,2-Trichloroethane	0.299	0.305	-2.0	60	0.00
56 T	Ethyl methacrylate	0.456	0.458	-0.4	59	0.00
57 T	1,3-Dichloropropane	0.515	0.508	1.4	58	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.201	5.6	58	0.00
59 T	2-Hexanone	0.239	0.249	-4.2	62	0.00
60 T	Dibromochloromethane	0.366	0.364	0.5	59	0.00
61 T	1,2-Dibromoethane	0.298	0.301	-1.0	60	0.00
62 S	4-Bromofluorobenzene	0.450	0.408	9.3	57	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.402	0.397	1.2	61	0.00
65 PM	Chlorobenzene	1.065	0.990	7.0	57	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.391	2.7	59	0.00
67 C	Ethyl Benzene	1.924	1.741	9.5#	55	0.00
68 T	m/p-Xylenes	0.734	0.665	9.4	56	0.00
69 T	o-Xylene	0.715	0.663	7.3	57	0.00
70 T	Styrene	1.206	1.142	5.3	58	0.00
71 P	Bromoform	0.272	0.265	2.6	59	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
73 T	Isopropylbenzene	3.915	3.435	12.3	57	0.00
74 T	N-amyl acetate	1.311	1.177	10.2	57	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.836	6.4	61	0.00
76 T	1,2,3-Trichloropropane	0.631	0.617	2.2	64	0.00
77 T	Bromobenzene	0.894	0.831	7.0	59	0.00
78 T	n-propylbenzene	4.718	4.063	13.9	55	0.00
79 T	2-Chlorotoluene	2.717	2.328	14.3	55	0.00
80 T	1,3,5-Trimethylbenzene	3.270	2.863	12.4	56	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.257	21.2	50#	0.00
82 T	4-Chlorotoluene	2.799	2.416	13.7	55	0.00
83 T	tert-Butylbenzene	2.900	2.534	12.6	55	0.00
84 T	1,2,4-Trimethylbenzene	3.209	2.844	11.4	56	0.00
85 T	sec-Butylbenzene	4.255	3.704	12.9	55	0.00
86 T	p-Isopropyltoluene	3.474	3.123	10.1	57	0.00
87 T	1,3-Dichlorobenzene	1.739	1.612	7.3	59	0.00
88 T	1,4-Dichlorobenzene	1.743	1.613	7.5	59	0.00
89 T	n-Butylbenzene	3.380	2.933	13.2	55	0.00
90 T	Hexachloroethane	0.714	0.627	12.2	56	0.00
91 T	1,2-Dichlorobenzene	1.594	1.502	5.8	60	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.162	5.3	62	0.00
93 T	1,2,4-Trichlorobenzene	1.012	0.969	4.2	59	0.00
94 T	Hexachlorobutadiene	0.556	0.508	8.6	56	0.00
95 T	Naphthalene	2.357	2.422	-2.8	64	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
Data File : VY014867.D
Acq On : 26 Jul 2023 22:09
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 27 01:15:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.883	1.8	61	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6