

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
 Data File : VY005468.D
 Acq On : 27 Jul 2021 16:38
 Operator : SY/MD
 Sample : M3142-02
 Misc : 5.74G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JONES-RD-STOCK-PILE-A

Quant Time: Jul 28 02:34:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	201514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	332603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	330361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	150128	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	131362	61.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.040%
35) Dibromofluoromethane	7.728	113	102741	53.498	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.000%
50) Toluene-d8	10.185	98	425937	52.328	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.660%
62) 4-Bromofluorobenzene	12.483	95	139005	51.743	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
Data File : VY005468.D
Acq On : 27 Jul 2021 16:38
Operator : SY/MD
Sample : M3142-02
Misc : 5.74G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JONES-RD-STOCK-PILE-A

Quant Time: Jul 28 02:34:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

