

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
 Data File : VY005475.D
 Acq On : 27 Jul 2021 19:20
 Operator : SY/MD
 Sample : M3165-02
 Misc : 5.70G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Jul 28 02:36:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	131544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	212153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	216252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	111104	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	83803	59.632	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.260%
35) Dibromofluoromethane	7.728	113	64065	52.298	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.600%
50) Toluene-d8	10.185	98	271206	52.236	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.480%
62) 4-Bromofluorobenzene	12.483	95	93409	54.511	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
Data File : VY005475.D
Acq On : 27 Jul 2021 19:20
Operator : SY/MD
Sample : M3165-02
Misc : 5.70G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Jul 28 02:36:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

