

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
 Data File : VY005477.D
 Acq On : 27 Jul 2021 20:06
 Operator : SY/MD
 Sample : M3200-06
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GS-3A

Quant Time: Jul 28 02:37:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	79533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	122226	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	107531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	50772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	37458	44.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.160%
35) Dibromofluoromethane	7.728	113	34306	48.610	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.220%
50) Toluene-d8	10.185	98	152095	50.847	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.700%
62) 4-Bromofluorobenzene	12.483	95	44127	44.698	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.400%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	19622	101.122	ug/l	96
25) 2-Butanone	7.003	43	3722	12.809	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072721\
Data File : VY005477.D
Acq On : 27 Jul 2021 20:06
Operator : SY/MD
Sample : M3200-06
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GS-3A

Quant Time: Jul 28 02:37:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

