

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
 Data File : VY009719.D
 Acq On : 27 Jul 2022 13:08
 Operator : KP/MD
 Sample : N3906-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-7-7-26-22

Quant Time: Jul 28 02:45:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	224660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	402326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	376371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	197838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	94721	64.880	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.760%
35) Dibromofluoromethane	7.710	113	98312	44.929	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.860%
50) Toluene-d8	10.179	98	365436	58.025	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	116.060%
62) 4-Bromofluorobenzene	12.477	95	118337	40.786	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	81.580%
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	2098	1.049	ug/l	99
16) Acetone	3.942	43	8255	15.368	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
Data File : VY009719.D
Acq On : 27 Jul 2022 13:08
Operator : KP/MD
Sample : N3906-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-7-7-26-22

Quant Time: Jul 28 02:45:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

