

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
 Data File : VY009720.D
 Acq On : 27 Jul 2022 13:31
 Operator : KP/MD
 Sample : N3839-05
 Misc : 5.69g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ENV-DE-WC-1

Quant Time: Jul 28 02:45:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	212850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	384749	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	355301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	172510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	96330	70.066	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.140%	
35) Dibromofluoromethane	7.710	113	94165	45.000	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 90.000%	
50) Toluene-d8	10.179	98	347913	57.746	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 115.500%	
62) 4-Bromofluorobenzene	12.477	95	116377	41.942	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 83.880%	
Target Compounds						
11) Tert butyl alcohol	4.954	59	2363	10.504	ug/l	Qvalue # 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
Data File : VY009720.D
Acq On : 27 Jul 2022 13:31
Operator : KP/MD
Sample : N3839-05
Misc : 5.69g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ENV-DE-WC-1

Quant Time: Jul 28 02:45:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

