

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
 Data File : VY009721.D
 Acq On : 27 Jul 2022 13:55
 Operator : KP/MD
 Sample : N3886-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WH-1

Quant Time: Jul 28 02:45:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	207234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	376926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	362612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	170213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	92880	69.331	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.660%	
35) Dibromofluoromethane	7.710	113	90928	44.355	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 88.720%	
50) Toluene-d8	10.173	98	335409	56.752	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 113.500%	
62) 4-Bromofluorobenzene	12.477	95	113705	41.830	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 83.660%	
Target Compounds						
25) 2-Butanone	6.984	43	1861	2.546	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
Data File : VY009721.D
Acq On : 27 Jul 2022 13:55
Operator : KP/MD
Sample : N3886-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WH-1

Quant Time: Jul 28 02:45:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

