

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
 Data File : VY009728.D
 Acq On : 27 Jul 2022 16:39
 Operator : KP/MD
 Sample : N3896-15
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-72

Quant Time: Jul 28 02:46:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	109973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	218520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	234197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	117282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	65756	94.417	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	188.840%	#
35) Dibromofluoromethane	7.709	113	59247	49.851	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	99.700%	
50) Toluene-d8	10.179	98	198543	58.044	ug/l	0.00
Spiked Amount	50.000	Range 78 - 125	Recovery	=	116.080%	
62) 4-Bromofluorobenzene	12.477	95	71764	45.539	ug/l	0.00
Spiked Amount	50.000	Range 50 - 146	Recovery	=	91.080%	
Target Compounds						
16) Acetone	3.942	43	6867	26.116	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
Data File : VY009728.D
Acq On : 27 Jul 2022 16:39
Operator : KP/MD
Sample : N3896-15
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-72

Quant Time: Jul 28 02:46:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

