

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
 Data File : VY009729.D
 Acq On : 27 Jul 2022 17:02
 Operator : KP/MD
 Sample : N3912-01
 Misc : 4.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 6601-S

Quant Time: Jul 28 02:46:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	181276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	331893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	330774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	163022	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	75961	64.447	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.900%
35) Dibromofluoromethane	7.710	113	77196	42.766	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.540%
50) Toluene-d8	10.173	98	290812	55.812	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	111.620%
62) 4-Bromofluorobenzene	12.477	95	95970	40.096	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	80.200%
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	1667	1.033	ug/l	92
39) Methylcyclohexane	9.173	83	3952	1.086	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072722\
Data File : VY009729.D
Acq On : 27 Jul 2022 17:02
Operator : KP/MD
Sample : N3912-01
Misc : 4.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
6601-S

Quant Time: Jul 28 02:46:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

