

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
 Data File : VY014870.D
 Acq On : 27 Jul 2023 10:23
 Operator : SY/MD
 Sample : VY0727SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0727SBL01

Quant Time: Jul 28 02:36:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	164291	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	289245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	274618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131171	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	101889	54.423	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.840%
35) Dibromofluoromethane	7.710	113	91987	49.959	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.920%
50) Toluene-d8	10.179	98	344408	50.735	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.477	95	126254	48.505	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
Data File : VY014870.D
Acq On : 27 Jul 2023 10:23
Operator : SY/MD
Sample : VY0727SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0727SBL01

Quant Time: Jul 28 02:36:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

