

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
 Data File : VY014875.D
 Acq On : 27 Jul 2023 12:20
 Operator : SY/MD
 Sample : 03686-10
 Misc : 3.34g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-29-2-3-20230721

Quant Time: Jul 28 02:37:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	155510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	276480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	264955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	126862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	100612	56.776	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.560%
35) Dibromofluoromethane	7.710	113	86873	49.360	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.720%
50) Toluene-d8	10.173	98	326165	50.266	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.540%
62) 4-Bromofluorobenzene	12.477	95	121739	48.930	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.860%
Target Compounds						
25) 2-Butanone	6.978	43	4158	5.061	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
Data File : VY014875.D
Acq On : 27 Jul 2023 12:20
Operator : SY/MD
Sample : 03686-10
Misc : 3.34g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-29-2-3-20230721

Quant Time: Jul 28 02:37:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

