

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
 Data File : VY014884.D
 Acq On : 27 Jul 2023 15:55
 Operator : SY/MD
 Sample : 03756-06
 Misc : 5.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPTP-10-GRAB-20-(4-5)

Quant Time: Jul 28 02:38:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	144458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	253043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	243836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	91570	55.627	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.260%
35) Dibromofluoromethane	7.710	113	78955	49.016	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.040%
50) Toluene-d8	10.173	98	299521	50.435	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.477	95	109364	48.027	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.060%
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	22087	9.624	ug/l	89
29) Tetrahydrofuran	7.344	42	1380	2.937	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
Data File : VY014884.D
Acq On : 27 Jul 2023 15:55
Operator : SY/MD
Sample : 03756-06
Misc : 5.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPTP-10-GRAB-20-(4-5)

Quant Time: Jul 28 02:38:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

