

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
 Data File : VY014869.D
 Acq On : 27 Jul 2023 09:24
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 02:35:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	60	0.00
2 T	Dichlorodifluoromethane	0.407	0.322	20.9	49#	0.00
3 P	Chloromethane	0.458	0.521	-13.8	69	0.00
4 C	Vinyl Chloride	0.512	0.589	-15.0#	68	0.00
5 T	Bromomethane	0.338	0.438	-29.6#	78	0.00
6 T	Chloroethane	0.320	0.400	-25.0#	73	0.00
7 T	Trichlorofluoromethane	0.860	0.841	2.2	57	0.00
8 T	Diethyl Ether	0.320	0.330	-3.1	61	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.496	5.9	55	0.00
10 T	Methyl Iodide	0.493	0.539	-9.3	59	0.00
11 T	Tert butyl alcohol	0.125	0.095	24.0	51	0.01
12 CM	1,1-Dichloroethene	0.492	0.468	4.9#	55	0.00
13 T	Acrolein	0.060	0.041	31.7#	44#	0.00
14 T	Allyl chloride	0.874	0.802	8.2	54	0.00
15 T	Acrylonitrile	0.176	0.191	-8.5	64	0.00
16 T	Acetone	0.181	0.174	3.9	53	0.00
17 T	Carbon Disulfide	1.410	1.189	15.7	51	0.00
18 T	Methyl Acetate	0.693	0.804	-16.0	68	0.00
19 T	Methyl tert-butyl Ether	1.599	1.719	-7.5	63	0.00
20 T	Methylene Chloride	0.656	0.630	4.0	62	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.544	2.7	57	0.00
22 T	Diisopropyl ether	1.777	1.796	-1.1	59	0.00
23 T	Vinyl Acetate	1.064	1.055	0.8	57	0.00
24 P	1,1-Dichloroethane	1.017	1.037	-2.0	59	0.00
25 T	2-Butanone	0.264	0.273	-3.4	60	0.00
26 T	2,2-Dichloropropane	0.969	0.957	1.2	58	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.681	-4.8	61	0.00
28 T	Bromochloromethane	0.438	0.413	5.7	59	0.00
29 T	Tetrahydrofuran	0.163	0.171	-4.9	63	0.00
30 C	Chloroform	1.049	1.107	-5.5#	62	0.00
31 T	Cyclohexane	0.946	0.763	19.3	50#	0.00
32 T	1,1,1-Trichloroethane	0.944	0.952	-0.8	59	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.524	8.1	56	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	62	0.00
35 S	Dibromofluoromethane	0.318	0.289	9.1	55	0.00
36 T	1,1-Dichloropropene	0.476	0.449	5.7	56	0.00
37 T	Ethyl Acetate	0.328	0.329	-0.3	63	0.00
38 T	Carbon Tetrachloride	0.503	0.494	1.8	58	0.00
39 T	Methylcyclohexane	0.620	0.530	14.5	51	0.00
40 TM	Benzene	1.354	1.359	-0.4	60	0.00
41 T	Methacrylonitrile	0.169	0.193	-14.2	71	0.00
42 TM	1,2-Dichloroethane	0.429	0.447	-4.2	62	0.00
43 T	Isopropyl Acetate	0.624	0.621	0.5	60	0.00
44 TM	Trichloroethene	0.379	0.376	0.8	60	0.00
45 C	1,2-Dichloropropane	0.353	0.355	-0.6#	60	0.00
46 T	Dibromomethane	0.214	0.225	-5.1	63	0.00
47 T	Bromodichloromethane	0.505	0.527	-4.4	62	0.00
48 T	Methyl methacrylate	0.281	0.281	0.0	60	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
 Data File : VY014869.D
 Acq On : 27 Jul 2023 09:24
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 02:35:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	65	0.00
50 S	Toluene-d8	1.173	1.025	12.6	54	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.348	-6.1	64	0.00
52 CM	Toluene	0.866	0.865	0.1#	60	0.00
53 T	t-1,3-Dichloropropene	0.553	0.558	-0.9	60	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.598	0.5	59	0.00
55 T	1,1,2-Trichloroethane	0.299	0.313	-4.7	62	0.00
56 T	Ethyl methacrylate	0.456	0.468	-2.6	60	0.00
57 T	1,3-Dichloropropane	0.515	0.533	-3.5	61	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.207	2.8	60	0.00
59 T	2-Hexanone	0.239	0.248	-3.8	62	0.00
60 T	Dibromochloromethane	0.366	0.390	-6.6	63	0.00
61 T	1,2-Dibromoethane	0.298	0.309	-3.7	62	0.00
62 S	4-Bromofluorobenzene	0.450	0.381	15.3	53	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	63	0.00
64 T	Tetrachloroethene	0.402	0.407	-1.2	63	0.00
65 PM	Chlorobenzene	1.065	1.062	0.3	61	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.418	-4.0	63	0.00
67 C	Ethyl Benzene	1.924	1.870	2.8#	59	0.00
68 T	m/p-Xylenes	0.734	0.716	2.5	60	0.00
69 T	o-Xylene	0.715	0.710	0.7	61	0.00
70 T	Styrene	1.206	1.212	-0.5	61	0.00
71 P	Bromoform	0.272	0.280	-2.9	62	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
73 T	Isopropylbenzene	3.915	3.714	5.1	60	0.00
74 T	N-amyl acetate	1.311	1.211	7.6	58	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.883	1.1	63	0.00
76 T	1,2,3-Trichloropropane	0.631	0.620	1.7	63	0.00
77 T	Bromobenzene	0.894	0.894	0.0	62	0.00
78 T	n-propylbenzene	4.718	4.389	7.0	58	0.00
79 T	2-Chlorotoluene	2.717	2.555	6.0	59	0.00
80 T	1,3,5-Trimethylbenzene	3.270	3.089	5.5	59	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.302	7.4	58	0.00
82 T	4-Chlorotoluene	2.799	2.642	5.6	59	0.00
83 T	tert-Butylbenzene	2.900	2.759	4.9	59	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.091	3.7	60	0.00
85 T	sec-Butylbenzene	4.255	4.019	5.5	59	0.00
86 T	p-Isopropyltoluene	3.474	3.320	4.4	59	0.00
87 T	1,3-Dichlorobenzene	1.739	1.747	-0.5	62	0.00
88 T	1,4-Dichlorobenzene	1.743	1.738	0.3	62	0.00
89 T	n-Butylbenzene	3.380	3.118	7.8	57	0.00
90 T	Hexachloroethane	0.714	0.666	6.7	58	0.00
91 T	1,2-Dichlorobenzene	1.594	1.618	-1.5	63	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.162	5.3	61	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.025	-1.3	61	0.00
94 T	Hexachlorobutadiene	0.556	0.528	5.0	58	0.00
95 T	Naphthalene	2.357	2.439	-3.5	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072723\
Data File : VY014869.D
Acq On : 27 Jul 2023 09:24
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 28 02:35:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.916	-1.9	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6