

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072821\
 Data File : VY005496.D
 Acq On : 28 Jul 2021 17:52
 Operator : SY/MD
 Sample : M3202-01
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-02-072721

Quant Time: Jul 29 05:34:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	14944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	21672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	18973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	7702	48.242	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.480%
35) Dibromofluoromethane	7.728	113	6352	50.761	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.520%
50) Toluene-d8	10.185	98	26452	49.874	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.740%
62) 4-Bromofluorobenzene	12.483	95	7303	41.720	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.440%
Target Compounds						
20) Methylene Chloride	4.722	84	1873	7.883	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072821\
Data File : VY005496.D
Acq On : 28 Jul 2021 17:52
Operator : SY/MD
Sample : M3202-01
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-02-072721

Quant Time: Jul 29 05:34:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

