

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072821\
 Data File : VY005480.D
 Acq On : 28 Jul 2021 11:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 29 05:29:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	0.453	0.369	18.5	110	0.00
3 P	Chloromethane	0.555	0.495	10.8	116	0.00
4 C	Vinyl Chloride	0.661	0.588	11.0#	116	0.00
5 T	Bromomethane	0.483	0.402	16.8	96	0.00
6 T	Chloroethane	0.415	0.380	8.4	107	0.00
7 T	Trichlorofluoromethane	0.864	0.786	9.0	108	0.00
8 T	Diethyl Ether	0.304	0.277	8.9	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.440	6.4	114	0.00
10 T	Methyl Iodide	0.533	0.512	3.9	115	0.00
11 T	Tert butyl alcohol	0.050	0.046	8.0	123	-0.05
12 CM	1,1-Dichloroethene	0.478	0.452	5.4#	113	0.00
13 T	Acrolein	0.046	0.041	10.9	127	0.00
14 T	Allyl chloride	0.855	0.791	7.5	110	0.00
15 T	Acrylonitrile	0.151	0.140	7.3	111	-0.01
16 T	Acetone	0.122	0.140	-14.8	157#	-0.02
17 T	Carbon Disulfide	1.581	1.419	10.2	109	0.00
18 T	Methyl Acetate	0.343	0.316	7.9	114	0.00
19 T	Methyl tert-butyl Ether	1.442	1.340	7.1	105	0.00
20 T	Methylene Chloride	0.707	0.547	22.6	100	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.495	7.8	108	0.00
22 T	Diisopropyl ether	1.660	1.632	1.7	109	0.00
23 T	Vinyl Acetate	1.092	1.077	1.4	107	0.00
24 P	1,1-Dichloroethane	0.915	0.913	0.2	111	0.00
25 T	2-Butanone	0.183	0.193	-5.5	125	0.00
26 T	2,2-Dichloropropane	0.832	0.827	0.6	114	0.00
27 T	cis-1,2-Dichloroethene	0.574	0.567	1.2	109	0.00
28 T	Bromochloromethane	0.368	0.371	-0.8	117	0.00
29 T	Tetrahydrofuran	0.121	0.121	0.0	113	0.00
30 C	Chloroform	0.905	0.908	-0.3#	107	0.00
31 T	Cyclohexane	0.912	0.883	3.2	114	0.00
32 T	1,1,1-Trichloroethane	0.838	0.828	1.2	107	0.00
33 S	1,2-Dichloroethane-d4	0.534	0.550	-3.0	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	115	0.00
35 S	Dibromofluoromethane	0.289	0.311	-7.6	112	0.00
36 T	1,1-Dichloropropene	0.483	0.494	-2.3	112	0.00
37 T	Ethyl Acetate	0.294	0.276	6.1	103	0.00
38 T	Carbon Tetrachloride	0.528	0.515	2.5	107	0.00
39 T	Methylcyclohexane	0.620	0.545	12.1	98	0.00
40 TM	Benzene	1.389	1.392	-0.2	109	0.00
41 T	Methacrylonitrile	0.154	0.159	-3.2	107	0.01
42 TM	1,2-Dichloroethane	0.449	0.432	3.8	103	0.00
43 T	Isopropyl Acetate	0.533	0.525	1.5	105	0.00
44 TM	Trichloroethene	0.377	0.371	1.6	107	0.00
45 C	1,2-Dichloropropane	0.342	0.293	14.3#	93	0.00
46 T	Dibromomethane	0.188	0.162	13.8	92	0.00
47 T	Bromodichloromethane	0.473	0.410	13.3	91	0.00
48 T	Methyl methacrylate	0.248	0.212	14.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072821\
 Data File : VY005480.D
 Acq On : 28 Jul 2021 11:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 29 05:29:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.002	33.3#	97	-0.03
50 S	Toluene-d8	1.224	1.135	7.3	97	0.00
51 T	4-Methyl-2-Pentanone	0.283	0.245	13.4	93	0.00
52 CM	Toluene	0.882	0.767	13.0#	93	0.00
53 T	t-1,3-Dichloropropene	0.529	0.453	14.4	91	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.497	14.6	91	0.00
55 T	1,1,2-Trichloroethane	0.270	0.233	13.7	92	0.00
56 T	Ethyl methacrylate	0.408	0.355	13.0	91	0.00
57 T	1,3-Dichloropropane	0.481	0.408	15.2	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.195	0.184	5.6	98	0.00
59 T	2-Hexanone	0.198	0.176	11.1	96	0.00
60 T	Dibromochloromethane	0.330	0.289	12.4	92	0.00
61 T	1,2-Dibromoethane	0.256	0.217	15.2	90	0.00
62 S	4-Bromofluorobenzene	0.404	0.379	6.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.378	0.384	-1.6	96	0.00
65 PM	Chlorobenzene	1.024	1.023	0.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.386	-1.0	92	0.00
67 C	Ethyl Benzene	1.864	1.888	-1.3#	94	0.00
68 T	m/p-Xylenes	0.724	0.735	-1.5	95	0.00
69 T	o-Xylene	0.689	0.698	-1.3	94	0.00
70 T	Styrene	1.169	1.192	-2.0	92	0.00
71 P	Bromoform	0.244	0.247	-1.2	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.448	3.486	-1.1	95	0.00
74 T	N-amyl acetate	1.050	1.031	1.8	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.658	0.640	2.7	91	0.00
76 T	1,2,3-Trichloropropane	0.492	0.492	0.0	96	0.00
77 T	Bromobenzene	0.844	0.839	0.6	93	0.00
78 T	n-propylbenzene	4.160	4.244	-2.0	97	0.00
79 T	2-Chlorotoluene	2.337	2.350	-0.6	95	0.00
80 T	1,3,5-Trimethylbenzene	2.880	2.941	-2.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.253	0.250	1.2	95	0.00
82 T	4-Chlorotoluene	2.418	2.435	-0.7	96	0.00
83 T	tert-Butylbenzene	2.565	2.633	-2.7	97	0.00
84 T	1,2,4-Trimethylbenzene	2.889	2.902	-0.4	95	0.00
85 T	sec-Butylbenzene	3.700	3.825	-3.4	98	0.00
86 T	p-Isopropyltoluene	3.186	3.314	-4.0	100	0.00
87 T	1,3-Dichlorobenzene	1.656	1.679	-1.4	95	0.00
88 T	1,4-Dichlorobenzene	1.641	1.657	-1.0	97	0.00
89 T	n-Butylbenzene	2.924	3.056	-4.5	101	0.00
90 T	Hexachloroethane	0.573	0.586	-2.3	97	0.00
91 T	1,2-Dichlorobenzene	1.479	1.481	-0.1	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.125	2.3	96	0.00
93 T	1,2,4-Trichlorobenzene	1.020	1.051	-3.0	100	0.00
94 T	Hexachlorobutadiene	0.647	0.685	-5.9	104	0.00
95 T	Naphthalene	2.052	2.049	0.1	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072821\
Data File : VY005480.D
Acq On : 28 Jul 2021 11:22
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 29 05:29:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.920	0.928	-0.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6