

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
 Data File : VY009746.D
 Acq On : 28 Jul 2022 13:17
 Operator : KP/MD
 Sample : N3894-09
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-28

Quant Time: Jul 29 02:56:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	154289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	276313	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	264402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	122407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	73423	73.970	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 147.940%	
35) Dibromofluoromethane	7.710	113	68317	45.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 90.920%	
50) Toluene-d8	10.179	98	245341	56.618	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 113.240%	
62) 4-Bromofluorobenzene	12.477	95	74473	37.373	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 74.740%	
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	14461	10.528	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
Data File : VY009746.D
Acq On : 28 Jul 2022 13:17
Operator : KP/MD
Sample : N3894-09
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-28

Quant Time: Jul 29 02:56:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

