

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
 Data File : VY009758.D
 Acq On : 28 Jul 2022 17:59
 Operator : KP/MD
 Sample : N3895-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-40

Quant Time: Jul 29 02:57:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	159187	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	286950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	281086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	130933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	64672	62.308	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.620%
35) Dibromofluoromethane	7.710	113	61902	39.664	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.320%
50) Toluene-d8	10.173	98	245798	54.457	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	108.920%
62) 4-Bromofluorobenzene	12.477	95	68353	33.031	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	66.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
Data File : VY009758.D
Acq On : 28 Jul 2022 17:59
Operator : KP/MD
Sample : N3895-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-40

Quant Time: Jul 29 02:57:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

