

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
 Data File : VY009763.D
 Acq On : 28 Jul 2022 19:57
 Operator : KP/MD
 Sample : N3881-03
 Misc : 9.76g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP2

Quant Time: Jul 29 02:58:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	179764	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	321717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	311674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	144491	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	69438	58.959	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.920%
35) Dibromofluoromethane	7.710	113	66302	37.893	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.780%
50) Toluene-d8	10.173	98	238068	46.414	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	92.820%
62) 4-Bromofluorobenzene	12.477	95	73480	31.671	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	63.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
Data File : VY009763.D
Acq On : 28 Jul 2022 19:57
Operator : KP/MD
Sample : N3881-03
Misc : 9.76g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP2

Quant Time: Jul 29 02:58:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

