

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
 Data File : VY009738.D
 Acq On : 28 Jul 2022 09:33
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 29 02:54:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.445	0.395	11.2	80	0.00
3 P	Chloromethane	0.545	0.532	2.4	83	0.00
4 C	Vinyl Chloride	0.608	0.645	-6.1#	90	0.00
5 T	Bromomethane	0.422	0.479	-13.5	96	0.00
6 T	Chloroethane	0.364	0.417	-14.6	95	0.00
7 T	Trichlorofluoromethane	0.803	0.755	6.0	81	0.00
8 T	Diethyl Ether	0.260	0.235	9.6	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.454	0.425	6.4	80	0.00
10 T	Methyl Iodide	0.590	0.587	0.5	79	0.00
11 T	Tert butyl alcohol	0.053	0.039	26.4#	73	0.00
12 CM	1,1-Dichloroethene	0.427	0.405	5.2#	80	0.00
13 T	Acrolein	0.009	0.006	33.3#	71	0.00
14 T	Allyl chloride	0.672	0.562	16.4	71	0.00
15 T	Acrylonitrile	0.125	0.114	8.8	77	0.00
16 T	Acetone	0.120	0.124	-3.3	75	0.00
17 T	Carbon Disulfide	1.314	1.215	7.5	76	0.00
18 T	Methyl Acetate	0.347	0.254	26.8#	73	0.00
19 T	Methyl tert-butyl Ether	1.166	1.119	4.0	79	0.00
20 T	Methylene Chloride	0.789	0.495	37.3#	74	0.00
21 T	trans-1,2-Dichloroethene	0.496	0.466	6.0	79	0.00
22 T	Diisopropyl ether	1.430	1.265	11.5	73	0.00
23 T	Vinyl Acetate	0.886	0.839	5.3	74	0.00
24 P	1,1-Dichloroethane	0.829	0.756	8.8	78	0.00
25 T	2-Butanone	0.176	0.166	5.7	73	0.00
26 T	2,2-Dichloropropane	0.788	0.719	8.8	80	0.00
27 T	cis-1,2-Dichloroethene	0.557	0.532	4.5	81	0.00
28 T	Bromochloromethane	0.299	0.279	6.7	80	0.00
29 T	Tetrahydrofuran	0.112	0.099	11.6	72	0.00
30 C	Chloroform	0.867	0.822	5.2#	81	0.00
31 T	Cyclohexane	0.796	0.660	17.1	72	0.00
32 T	1,1,1-Trichloroethane	0.778	0.754	3.1	82	0.00
33 S	1,2-Dichloroethane-d4	0.384	0.317	17.4	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.272	0.252	7.4	84	0.00
36 T	1,1-Dichloropropene	0.442	0.424	4.1	77	0.00
37 T	Ethyl Acetate	0.262	0.234	10.7	73	0.00
38 T	Carbon Tetrachloride	0.474	0.482	-1.7	83	0.00
39 T	Methylcyclohexane	0.548	0.524	4.4	75	0.00
40 TM	Benzene	1.311	1.250	4.7	78	0.00
41 T	Methacrylonitrile	0.134	0.143	-6.7	86	0.02
42 TM	1,2-Dichloroethane	0.366	0.367	-0.3	82	0.00
43 T	Isopropyl Acetate	0.458	0.431	5.9	75	0.00
44 TM	Trichloroethene	0.376	0.369	1.9	81	0.00
45 C	1,2-Dichloropropane	0.318	0.297	6.6#	76	0.00
46 T	Dibromomethane	0.189	0.190	-0.5	81	0.00
47 T	Bromodichloromethane	0.435	0.441	-1.4	81	0.00
48 T	Methyl methacrylate	0.201	0.191	5.0	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
 Data File : VY009738.D
 Acq On : 28 Jul 2022 09:33
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 29 02:54:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	94	0.00
50 S	Toluene-d8	0.911	0.752	17.5	88	0.00
51 T	4-Methyl-2-Pentanone	0.233	0.231	0.9	83	0.00
52 CM	Toluene	0.821	0.818	0.4#	85	0.00
53 T	t-1,3-Dichloropropene	0.449	0.455	-1.3	85	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.504	2.3	77	0.00
55 T	1,1,2-Trichloroethane	0.268	0.267	0.4	87	0.00
56 T	Ethyl methacrylate	0.334	0.347	-3.9	84	0.00
57 T	1,3-Dichloropropane	0.436	0.427	2.1	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.163	-12.4	78	0.00
59 T	2-Hexanone	0.166	0.172	-3.6	82	0.00
60 T	Dibromochloromethane	0.324	0.335	-3.4	90	0.00
61 T	1,2-Dibromoethane	0.261	0.261	0.0	86	0.00
62 S	4-Bromofluorobenzene	0.361	0.353	2.2	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.394	0.397	-0.8	85	0.00
65 PM	Chlorobenzene	1.003	0.968	3.5	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.378	-1.3	83	0.00
67 C	Ethyl Benzene	1.734	1.687	2.7#	78	0.00
68 T	m/p-Xylenes	0.688	0.689	-0.1	80	0.00
69 T	o-Xylene	0.644	0.648	-0.6	80	0.00
70 T	Styrene	1.082	1.101	-1.8	80	0.00
71 P	Bromoform	0.238	0.248	-4.2	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.324	3.177	4.4	79	0.00
74 T	N-amyl acetate	0.859	0.806	6.2	74	0.00
75 P	1,1,2,2-Tetrachloroethane	0.654	0.616	5.8	78	0.00
76 T	1,2,3-Trichloropropane	0.464	0.463	0.2	82	0.00
77 T	Bromobenzene	0.825	0.809	1.9	82	0.00
78 T	n-propylbenzene	3.950	3.752	5.0	78	0.00
79 T	2-Chlorotoluene	2.235	2.082	6.8	77	0.00
80 T	1,3,5-Trimethylbenzene	2.755	2.649	3.8	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.233	0.217	6.9	76	0.00
82 T	4-Chlorotoluene	2.325	2.187	5.9	79	0.00
83 T	tert-Butylbenzene	2.476	2.324	6.1	79	0.00
84 T	1,2,4-Trimethylbenzene	2.777	2.637	5.0	79	0.00
85 T	sec-Butylbenzene	3.635	3.417	6.0	78	0.00
86 T	p-Isopropyltoluene	3.074	2.946	4.2	79	0.00
87 T	1,3-Dichlorobenzene	1.667	1.585	4.9	82	0.00
88 T	1,4-Dichlorobenzene	1.678	1.576	6.1	82	0.00
89 T	n-Butylbenzene	2.816	2.617	7.1	77	0.00
90 T	Hexachloroethane	0.605	0.547	9.6	79	0.00
91 T	1,2-Dichlorobenzene	1.489	1.417	4.8	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.110	0.099	10.0	77	0.00
93 T	1,2,4-Trichlorobenzene	0.939	0.910	3.1	82	0.00
94 T	Hexachlorobutadiene	0.569	0.521	8.4	81	0.00
95 T	Naphthalene	1.761	1.843	-4.7	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072822\
Data File : VY009738.D
Acq On : 28 Jul 2022 09:33
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 29 02:54:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.813	1.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6