

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072823\
 Data File : VY014901.D
 Acq On : 28 Jul 2023 12:08
 Operator : SY/MD
 Sample : 03756-19
 Misc : 4.87g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 307/308-TP-6-8-GRAB-(4-5)

Quant Time: Jul 29 03:02:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	141601	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	252361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	241738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	117463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	92921	57.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.180%
35) Dibromofluoromethane	7.710	113	80214	49.932	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.860%
50) Toluene-d8	10.173	98	300331	50.708	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.420%
62) 4-Bromofluorobenzene	12.477	95	108910	47.957	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.920%
Target Compounds						
20) Methylene Chloride	4.710	84	26245	12.645	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072823\
Data File : VY014901.D
Acq On : 28 Jul 2023 12:08
Operator : SY/MD
Sample : 03756-19
Misc : 4.87g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
307/308-TP-6-8-GRAB-(4-5)

Quant Time: Jul 29 03:02:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

