

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072823\
 Data File : VY014904.D
 Acq On : 28 Jul 2023 13:19
 Operator : SY/MD
 Sample : 03756-23
 Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 415/417-TP-4-(3-4)

Quant Time: Jul 29 03:03:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	138617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	243586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	239896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	91723	58.067	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.140%	
35) Dibromofluoromethane	7.710	113	78192	50.427	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.860%	
50) Toluene-d8	10.173	98	293011	51.255	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.500%	
62) 4-Bromofluorobenzene	12.477	95	109546	49.975	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.940%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	20487	9.149	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072823\
Data File : VY014904.D
Acq On : 28 Jul 2023 13:19
Operator : SY/MD
Sample : 03756-23
Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
415/417-TP-4-(3-4)

Quant Time: Jul 29 03:03:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

