

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009767.D
 Acq On : 29 Jul 2022 13:01
 Operator : KP/MD
 Sample : VY0729SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0729SBL01

Quant Time: Aug 01 09:14:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	164396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	283998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	284810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	120836	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	70005	65.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.180%
35) Dibromofluoromethane	7.709	113	68100	44.089	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.180%
50) Toluene-d8	10.178	98	262015	59.011	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	118.020%
62) 4-Bromofluorobenzene	12.483	95	75688	36.955	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	73.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
Data File : VY009767.D
Acq On : 29 Jul 2022 13:01
Operator : KP/MD
Sample : VY0729SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBL01

Quant Time: Aug 01 09:14:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

