

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009770.D
 Acq On : 29 Jul 2022 14:18
 Operator : KP/MD
 Sample : N3930-08
 Misc : 2.49g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-100

Quant Time: Aug 01 01:12:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	148965	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	258879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	253530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	118179	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	63412	65.561	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.120%
35) Dibromofluoromethane	7.710	113	61231	43.489	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.980%
50) Toluene-d8	10.179	98	225010	55.325	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	110.660%
62) 4-Bromofluorobenzene	12.477	95	70032	37.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	75.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
Data File : VY009770.D
Acq On : 29 Jul 2022 14:18
Operator : KP/MD
Sample : N3930-08
Misc : 2.49g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-100

Quant Time: Aug 01 01:12:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

