

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009775.D
 Acq On : 29 Jul 2022 16:16
 Operator : KP/MD
 Sample : N3923-01RE
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-03-072722RE

Quant Time: Aug 01 01:14:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	3422	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	5374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1726	78.746	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.500%
35) Dibromofluoromethane	7.710	113	510	17.449	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	34.900%#
50) Toluene-d8	10.173	98	4787	56.815	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	113.640%
62) 4-Bromofluorobenzene	12.477	95	1305	33.673	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	67.340%
Target Compounds						
20) Methylene Chloride	4.704	84	1034	20.639	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
Data File : VY009775.D
Acq On : 29 Jul 2022 16:16
Operator : KP/MD
Sample : N3923-01RE
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-03-072722RE

Quant Time: Aug 01 01:14:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

