

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009780.D
 Acq On : 29 Jul 2022 18:13
 Operator : KP/MD
 Sample : N3927-11
 Misc : 4.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-84

Quant Time: Aug 01 01:15:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	154372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	272787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	278648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	128158	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	62365	61.927	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.860%
35) Dibromofluoromethane	7.710	113	58989	39.760	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.520%
50) Toluene-d8	10.173	98	240055	56.073	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	112.140%
62) 4-Bromofluorobenzene	12.477	95	67035	34.076	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	68.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
Data File : VY009780.D
Acq On : 29 Jul 2022 18:13
Operator : KP/MD
Sample : N3927-11
Misc : 4.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-84

Quant Time: Aug 01 01:15:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

