

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009782.D
 Acq On : 29 Jul 2022 19:00
 Operator : KP/MD
 Sample : N3927-19
 Misc : 3.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-92

Quant Time: Aug 01 01:16:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1829	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	3076	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	834	70.638	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.280%
35) Dibromofluoromethane	7.710	113	288	17.215	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	34.440%#
50) Toluene-d8	10.173	98	2809	58.363	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	116.720%
62) 4-Bromofluorobenzene	12.477	95	573	25.830	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	51.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009782.D
 Acq On : 29 Jul 2022 19:00
 Operator : KP/MD
 Sample : N3927-19
 Misc : 3.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-92

Quant Time: Aug 01 01:16:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

