

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009768.D
 Acq On : 29 Jul 2022 13:31
 Operator : KP/MD
 Sample : VY0729SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0729SBS01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/01/2022
 Supervised By :Mahesh Dadoda 08/01/2022

Quant Time: Aug 01 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	165440	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	248365	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	248218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	123252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	54537	49.473	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.940%		
35) Dibromofluoromethane	7.710	113	62005	45.903	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	91.800%		
50) Toluene-d8	10.173	98	198418	50.477	ug/l	0.00
Spiked Amount 50.000	Range 78 - 125		Recovery =	100.960%		
62) 4-Bromofluorobenzene	12.477	95	86343	48.206	ug/l	0.00
Spiked Amount 50.000	Range 50 - 146		Recovery =	96.420%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.900	85	30117	20.448	ug/l	98
3) Chloromethane	2.107	50	43003	23.847	ug/l	99
4) Vinyl Chloride	2.241	62	46879	23.295	ug/l	98
5) Bromomethane	2.631	94	34490	24.722	ug/l	94
6) Chloroethane	2.778	64	30583	25.404	ug/l	99
7) Trichlorofluoromethane	3.113	101	55609	20.918	ug/l	96
8) Diethyl Ether	3.521	74	17094	19.838	ug/l	84
9) 1,1,2-Trichlorotrifluo...	3.881	101	30583	20.344	ug/l	95
10) Methyl Iodide	4.076	142	36336	18.629	ug/l	91
11) Tert butyl alcohol	4.948	59	18543	113.487	ug/l	99
12) 1,1-Dichloroethene	3.857	96	28151	19.919	ug/l	83
13) Acrolein	3.717	56	3548	130.453	ug/l	100
14) Allyl chloride	4.466	41	43101	19.397	ug/l	# 92
15) Acrylonitrile	5.149	53	43304	104.347	ug/l	98
16) Acetone	3.936	43	37642	95.161	ug/l	91
17) Carbon Disulfide	4.180	76	87025	20.013	ug/l	98
18) Methyl Acetate	4.466	43	22073	20.941	ug/l	93
19) Methyl tert-butyl Ether	5.210	73	74580	19.324	ug/l	98
20) Methylene Chloride	4.704	84	52273	22.131	ug/l	# 81
21) trans-1,2-Dichloroethene	5.204	96	32441	19.755	ug/l	88
22) Diisopropyl ether	6.106	45	96888	20.474	ug/l	94
23) Vinyl Acetate	6.051	43	301228	99.263	ug/l	# 92
24) 1,1-Dichloroethane	6.009	63	56125	20.465	ug/l	97
25) 2-Butanone	6.978	43	58201	99.739	ug/l	# 89
26) 2,2-Dichloropropane	6.972	77	52773	20.242	ug/l	98
27) cis-1,2-Dichloroethene	6.978	96	35741	19.402	ug/l	91
28) Bromochloromethane	7.326	49	19826	20.054	ug/l	# 86
29) Tetrahydrofuran	7.338	42	37619	101.802	ug/l	91
30) Chloroform	7.496	83	58546	20.406	ug/l	99
31) Cyclohexane	7.771	56	50037	19.009	ug/l	92
32) 1,1,1-Trichloroethane	7.691	97	51596	20.035	ug/l	97
36) 1,1-Dichloropropene	7.911	75	44312	20.160	ug/l	97
37) Ethyl Acetate	7.070	43	26762	20.543	ug/l	# 95
38) Carbon Tetrachloride	7.893	117	48727	20.689	ug/l	96
39) Methylcyclohexane	9.179	83	51203	18.806	ug/l	92
40) Benzene	8.155	78	132347	20.322	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009768.D
 Acq On : 29 Jul 2022 13:31
 Operator : KP/MD
 Sample : VY0729SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0729SBS01

Quant Time: Aug 01 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/01/2022
 Supervised By :Mahesh Dadoda 08/01/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.301	41	13251m	19.965	ug/l	
42) 1,2-Dichloroethane	8.228	62	37355	20.573	ug/l	91
43) Isopropyl Acetate	8.271	43	47087	20.689	ug/l	94
44) Trichloroethene	8.935	130	36801	19.698	ug/l	95
45) 1,2-Dichloropropane	9.209	63	32771	20.725	ug/l	93
46) Dibromomethane	9.301	93	19207	20.417	ug/l	93
47) Bromodichloromethane	9.490	83	46956	21.727	ug/l	98
48) Methyl methacrylate	9.289	41	20109	20.092	ug/l #	87
49) 1,4-Dioxane	9.289	88	4952	412.253	ug/l	88
51) 4-Methyl-2-Pentanone	10.063	43	127719	110.308	ug/l	91
52) Toluene	10.240	92	87952	21.561	ug/l	97
53) t-1,3-Dichloropropene	10.459	75	48058	21.532	ug/l	96
54) cis-1,3-Dichloropropene	9.923	75	53868	21.015	ug/l #	85
55) 1,1,2-Trichloroethane	10.642	97	28981	21.788	ug/l	96
56) Ethyl methacrylate	10.508	69	33731	20.355	ug/l	85
57) 1,3-Dichloropropane	10.788	76	47945	22.163	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.776	63	54878	82.341	ug/l	92
59) 2-Hexanone	10.831	43	87758	106.352	ug/l	92
60) Dibromochloromethane	10.984	129	35683	22.180	ug/l	97
61) 1,2-Dibromoethane	11.087	107	27472	21.203	ug/l	96
64) Tetrachloroethene	10.715	164	38013	19.423	ug/l	97
65) Chlorobenzene	11.514	112	98287	19.733	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.587	131	36617	19.755	ug/l	98
67) Ethyl Benzene	11.593	91	163274	18.965	ug/l	96
68) m/p-Xylenes	11.697	106	132825	38.865	ug/l	90
69) o-Xylene	12.026	106	58030	18.157	ug/l	91
70) Styrene	12.044	104	100260	18.665	ug/l	93
71) Bromoform	12.203	173	22770	19.238	ug/l #	99
73) Isopropylbenzene	12.325	105	154099	18.808	ug/l	100
74) N-amyl acetate	12.142	43	40642	19.204	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.581	83	32678	20.277	ug/l	98
76) 1,2,3-Trichloropropane	12.630	75	22965m	20.060	ug/l	
77) Bromobenzene	12.605	156	40177	19.745	ug/l	88
78) n-propylbenzene	12.666	91	188035	19.312	ug/l	97
79) 2-Chlorotoluene	12.752	91	103816	18.840	ug/l	93
80) 1,3,5-Trimethylbenzene	12.812	105	130333	19.194	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.374	75	11146	19.407	ug/l	87
82) 4-Chlorotoluene	12.849	91	109588	19.117	ug/l	95
83) tert-Butylbenzene	13.075	119	112800	18.485	ug/l	91
84) 1,2,4-Trimethylbenzene	13.117	105	129138	18.868	ug/l	96
85) sec-Butylbenzene	13.251	105	171795	19.174	ug/l	98
86) p-Isopropyltoluene	13.367	119	142418	18.794	ug/l	96
87) 1,3-Dichlorobenzene	13.361	146	79726	19.405	ug/l	98
88) 1,4-Dichlorobenzene	13.440	146	80088	19.357	ug/l	98
89) n-Butylbenzene	13.696	91	131143	18.890	ug/l	99
90) Hexachloroethane	13.959	117	28035	18.807	ug/l	98
91) 1,2-Dichlorobenzene	13.739	146	71479	19.480	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.355	75	5028	18.567	ug/l	72
93) 1,2,4-Trichlorobenzene	15.007	180	42396	18.307	ug/l	99
94) Hexachlorobutadiene	15.111	225	26630	18.992	ug/l	99
95) Naphthalene	15.233	128	77396	17.829	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	37776	18.650	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009768.D
 Acq On : 29 Jul 2022 13:31
 Operator : KP/MD
 Sample : VY0729SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 VY0729SBS01

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/01/2022
 Supervised By :Mahesh Dadoda 08/01/2022

Quant Time: Aug 01 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009768.D
 Acq On : 29 Jul 2022 13:31
 Operator : KP/MD
 Sample : VY0729SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0729SBS01

Quant Time: Aug 01 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Krupa Patel 08/01/2022
 Supervised By :Mahesh Dadoda 08/01/2022





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:168 Resp: 165440
Ion Ratio Lower Upper
168 100
99 45.5 46.9 70.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#2
Dichlorodifluoromethane
Concen: 20.448 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 85 Resp: 30117
Ion Ratio Lower Upper
85 100
87 33.2 16.1 48.2





#3
Chloromethane
Concen: 23.847 ug/l
RT: 2.107 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 50 Resp: 43000
Ion Ratio Lower Upper
50 100
52 31.6 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#4
Vinyl Chloride
Concen: 23.295 ug/l
RT: 2.241 min Scan# 69
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 62 Resp: 46879
Ion Ratio Lower Upper
62 100
64 32.7 25.0 37.6





#5
Bromomethane
Concen: 24.722 ug/l
RT: 2.631 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 94 Resp: 34490
Ion Ratio Lower Upper
94 100
96 86.6 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#6
Chloroethane
Concen: 25.404 ug/l
RT: 2.778 min Scan# 157
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 64 Resp: 30583
Ion Ratio Lower Upper
64 100
66 32.2 25.4 38.0





#7
Trichlorofluoromethane
Concen: 20.918 ug/l
RT: 3.113 min Scan# 211
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:101 Resp: 55609
Ion Ratio Lower Upper
101 100
103 62.8 48.0 72.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#8
Diethyl Ether
Concen: 19.838 ug/l
RT: 3.521 min Scan# 279
Delta R.T. 0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 74 Resp: 17094
Ion Ratio Lower Upper
74 100
45 92.9 55.1 165.5





#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.344 ug/l
RT: 3.881 min Scan# 31
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion:101 Resp: 3058
Ion Ratio Lower Upper
101 100
85 43.7 40.6 60.8
151 83.4 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#10
Methyl Iodide
Concen: 18.629 ug/l
RT: 4.076 min Scan# 370
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion:142 Resp: 36336
Ion Ratio Lower Upper
142 100
127 43.5 41.5 62.3
141 14.7 11.9 17.9





#11
Tert butyl alcohol
Concen: 113.487 ug/l
RT: 4.948 min Scan# 514
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 59 Resp: 1854
Ion Ratio Lower Upper
59 100
57 8.4 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#12
1,1-Dichloroethene
Concen: 19.919 ug/l
RT: 3.857 min Scan# 334
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 96 Resp: 28151
Ion Ratio Lower Upper
96 100
61 151.1 145.8 218.6
98 63.4 52.5 78.7





#13
Acrolein
Concen: 130.453 ug/l
RT: 3.717 min Scan# 312
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 56 Resp: 3548
Ion Ratio Lower Upper
56 100
55 72.2 57.6 86.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#14
Allyl chloride
Concen: 19.397 ug/l
RT: 4.466 min Scan# 434
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 41 Resp: 43101
Ion Ratio Lower Upper
41 100
39 68.1 58.8 88.2
76 38.4 25.1 37.7#





#15
Acrylonitrile
Concen: 104.347 ug/l
RT: 5.149 min Scan# 546
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 53 Resp: 43304
Ion Ratio Lower Upper
53 100
52 81.6 67.0 100.4
51 36.8 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#16
Acetone
Concen: 95.161 ug/l
RT: 3.936 min Scan# 347
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 43 Resp: 37642
Ion Ratio Lower Upper
43 100
58 32.2 22.0 33.0





#17
Carbon Disulfide
Concen: 20.013 ug/l
RT: 4.180 min Scan# 387
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 76 Resp: 87025
Ion Ratio Lower Upper
76 100
78 9.8 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#18
Methyl Acetate
Concen: 20.941 ug/l
RT: 4.466 min Scan# 434
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion: 43 Resp: 22073
Ion Ratio Lower Upper
43 100
74 24.0 16.6 24.8









#19
Methyl tert-butyl Ether
Concen: 19.324 ug/l
RT: 5.210 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01

Tgt Ion: 73 Resp: 74580
Ion Ratio Lower Upper
73 100
57 22.1 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#20
Methylene Chloride
Concen: 22.131 ug/l
RT: 4.704 min Scan# 473
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 84 Resp: 52273
Ion Ratio Lower Upper
84 100
49 110.2 114.4 171.6#
51 35.2 35.3 52.9#
86 60.8 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 19.755 ug/l
RT: 5.204 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 96 Resp: 3244
Ion Ratio Lower Upper
96 100
61 129.1 120.5 180.7
98 62.8 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#22
Diisopropyl ether
Concen: 20.474 ug/l
RT: 6.106 min Scan# 703
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 45 Resp: 96888
Ion Ratio Lower Upper
45 100
43 57.8 49.3 73.9
87 28.4 19.2 28.8
59 12.6 8.4 12.6





#23
 Vinyl Acetate
 Concen: 99.263 ug/l
 RT: 6.051 min Scan# 694
 Delta R.T. 0.000 min
 Lab File: VY009768.D
 Acq: 29 Jul 2022 13:31




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0729SBS01




Tgt Ion: 43 Resp: 301228
 Ion Ratio Lower Upper
 43 100
 86 11.9 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 08/01/2022
 Supervised By :Mahesh Dadoda 08/01/2022




#24
 1,1-Dichloroethane
 Concen: 20.465 ug/l
 RT: 6.009 min Scan# 687
 Delta R.T. 0.006 min
 Lab File: VY009768.D
 Acq: 29 Jul 2022 13:31




Tgt Ion: 63 Resp: 56125
 Ion Ratio Lower Upper
 63 100
 98 8.2 3.6 10.8
 100 4.9 2.0 5.8





#25
2-Butanone
Concen: 99.739 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 43 Resp: 5820
Ion Ratio Lower Upper
43 100
72 27.4 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#26
2,2-Dichloropropane
Concen: 20.242 ug/l
RT: 6.972 min Scan# 845
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 77 Resp: 52773
Ion Ratio Lower Upper
77 100
97 22.0 10.4 31.4





#27
cis-1,2-Dichloroethene
Concen: 19.402 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 96 Resp: 3574
Ion Ratio Lower Upper
96 100
61 144.2 0.0 319.8
98 65.1 0.0 127.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#28
Bromochloromethane
Concen: 20.054 ug/l
RT: 7.326 min Scan# 903
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion: 49 Resp: 19826
Ion Ratio Lower Upper
49 100
129 2.4 0.0 2.2#
130 80.4 55.4 83.0





#29

Tetrahydrofuran

Concen: 101.802 ug/l

RT: 7.338 min Scan# 905

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01





Tgt Ion:	42	Resp:	37619
Ion Ratio	Lower	Upper	
42	100		
72	44.1	30.3	45.5
71	41.0	29.0	43.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#30

Chloroform

Concen: 20.406 ug/l




RT: 7.496 min Scan# 931


Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31


Tgt Ion:	83	Resp:	58546
Ion Ratio	Lower	Upper	
83	100		
85	65.8	51.8	77.6





#31
Cyclohexane
Concen: 19.009 ug/l
RT: 7.771 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 56 Resp: 5003
Ion Ratio Lower Upper
56 100
69 37.3 26.7 40.1
84 91.6 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#32
1,1,1-Trichloroethane
Concen: 20.035 ug/l
RT: 7.691 min Scan# 963
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion: 97 Resp: 51596
Ion Ratio Lower Upper
97 100
99 64.2 50.7 76.1
61 42.4 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 49.473 ug/l

RT: 8.136 min Scan# 1036

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion: 65 Resp: 5453

Ion Ratio Lower Upper

65 100





67 58.7 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31




Tgt Ion:114 Resp: 248365


Ion Ratio Lower Upper

114 100

63 18.1 0.0 43.2


88 14.1 0.0 31.8





#35
Dibromofluoromethane
Concen: 45.903 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 113 Resp: 6200
Ion Ratio Lower Upper
113 100
111 102.4 82.6 124.0
192 22.3 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#36
1,1-Dichloropropene
Concen: 20.160 ug/l
RT: 7.911 min Scan# 999
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 75 Resp: 44312
Ion Ratio Lower Upper
75 100
110 36.8 17.1 51.2
77 31.1 24.4 36.6





#37
Ethyl Acetate
Concen: 20.543 ug/l
RT: 7.070 min Scan# 861
Delta R.T. 0.007 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 43 Resp: 26762
Ion Ratio Lower Upper
43 100
61 14.7 9.7 14.5
70 10.0 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#38
Carbon Tetrachloride
Concen: 20.689 ug/l
RT: 7.893 min Scan# 996
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion:117 Resp: 48727
Ion Ratio Lower Upper
117 100
119 92.7 77.9 116.9
121 32.0 24.8 37.2





#39

Methylcyclohexane

Concen: 18.806 ug/l

RT: 9.179 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01





Tgt Ion: 83	Resp: 5120
Ion Ratio	Lower Upper
83 100	
55 73.8	66.5 99.7
98 50.8	37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#40

Benzene

Concen: 20.322 ug/l




RT: 8.155 min Scan# 1039


Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31


Tgt Ion: 78	Resp: 132347
Ion Ratio	Lower Upper
78 100	
77 24.1	20.4 30.6





#41
Methacrylonitrile
Concen: 19.965 ug/l m
RT: 7.301 min Scan# 899
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 41 Resp: 1325
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#42
1,2-Dichloroethane
Concen: 20.573 ug/l
RT: 8.228 min Scan# 1051
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 62 Resp: 37355
Ion Ratio Lower Upper
62 100
98 10.7 0.0 14.8





#43
Isopropyl Acetate
Concen: 20.689 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.007 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 43 Resp: 4708
Ion Ratio Lower Upper
43 100
61 28.6 20.2 30.2
87 13.2 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#44
Trichloroethene
Concen: 19.698 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion:130 Resp: 36801
Ion Ratio Lower Upper
130 100
95 89.5 0.0 188.2





#45

1,2-Dichloropropane

Concen: 20.725 ug/l

RT: 9.209 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion: 63 Resp: 3277

Ion Ratio Lower Upper

63 100





65 32.6 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#46

Dibromomethane

Concen: 20.417 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31




Tgt Ion: 93 Resp: 19207


Ion Ratio Lower Upper

93 100

95 86.2 64.3 96.5


174 112.3 83.4 125.0





#47
Bromodichloromethane
Concen: 21.727 ug/l
RT: 9.490 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 83 Resp: 46950
Ion Ratio Lower Upper
83 100
85 66.5 51.9 77.9
127 9.8 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#48
Methyl methacrylate
Concen: 20.092 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion: 41 Resp: 20109
Ion Ratio Lower Upper
41 100
69 89.4 58.2 87.2#
39 56.6 47.4 71.0





#49

1,4-Dioxane

Concen: 412.253 ug/l

RT: 9.289 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01





Tgt Ion: 88	Resp: 495
Ion Ratio	Lower Upper
88	100
43	36.0 32.8 49.2
58	62.5 59.8 89.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#50

Toluene-d8

Concen: 50.477 ug/l




RT: 10.173 min Scan# 1370


Delta R.T. -0.006 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31


Tgt Ion: 98	Resp: 198418
Ion Ratio	Lower Upper
98	100
100	64.8 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 110.308 ug/l
RT: 10.063 min Scan# 1352
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 43 Resp: 127719
Ion Ratio Lower Upper
43 100
58 40.3 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#52
Toluene
Concen: 21.561 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 92 Resp: 87952
Ion Ratio Lower Upper
92 100
91 174.2 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 21.532 ug/l
RT: 10.459 min Scan# 1417
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 75 Resp: 48058
Ion Ratio Lower Upper
75 100
77 29.8 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#54
cis-1,3-Dichloropropene
Concen: 21.015 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 75 Resp: 53868
Ion Ratio Lower Upper
75 100
77 31.9 25.6 38.4
39 42.5 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 21.788 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 97	Resp: 2898
Ion Ratio	Lower Upper
97 100	
83 82.2	69.8 104.8
85 52.4	44.3 66.5
99 63.4	49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#56
Ethyl methacrylate
Concen: 20.355 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion: 69	Resp: 33731
Ion Ratio	Lower Upper
69 100	
41 65.4	63.4 95.0
39 36.3	35.8 53.8





#57

1,3-Dichloropropane

Concen: 22.163 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion: 76 Resp: 4794

Ion Ratio Lower Upper

76 100





78 32.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#58

2-Chloroethyl Vinyl ether

Concen: 82.341 ug/l

RT: 9.776 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VY009768.D




Acq: 29 Jul 2022 13:31


Tgt Ion: 63 Resp: 54878

Ion Ratio Lower Upper

63 100


106 31.3 21.5 32.3





#59
2-Hexanone
Concen: 106.352 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 43 Resp: 87758
Ion Ratio Lower Upper
43 100
58 54.6 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#60
Dibromochloromethane
Concen: 22.180 ug/l
RT: 10.984 min Scan# 1503
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion:129 Resp: 35683
Ion Ratio Lower Upper
129 100
127 74.7 38.7 116.1





#61
1,2-Dibromoethane
Concen: 21.203 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion:107 Resp: 27472
Ion Ratio Lower Upper
107 100
109 98.3 75.4 113.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#62
4-Bromofluorobenzene
Concen: 48.206 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion: 95 Resp: 86343
Ion Ratio Lower Upper
95 100
174 96.6 0.0 177.0
176 92.7 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion:117 Resp: 248218
Ion Ratio Lower Upper
117 100
82 50.5 43.0 64.4
119 33.0 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#64
Tetrachloroethene
Concen: 19.423 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion:164 Resp: 38013
Ion Ratio Lower Upper
164 100
166 131.2 100.4 150.6
129 87.8 71.4 107.0
131 87.6 70.8 106.2









#65
Chlorobenzene
Concen: 19.733 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01

Tgt Ion:112 Resp: 9828
Ion Ratio Lower Upper
112 100
114 31.3 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 19.755 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion:131 Resp: 36617
Ion Ratio Lower Upper
131 100
133 95.3 47.9 143.6
119 64.5 30.8 92.3





#67

Ethyl Benzene

Concen: 18.965 ug/l

RT: 11.593 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion: 91 Resp: 163274

Ion Ratio Lower Upper

91 100





106 32.8 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#68

m/p-Xylenes

Concen: 38.865 ug/l

RT: 11.697 min Scan# 1620

Delta R.T. -0.000 min

Lab File: VY009768.D




Acq: 29 Jul 2022 13:31


Tgt Ion:106 Resp: 132825

Ion Ratio Lower Upper

106 100


91 192.8 166.6 249.8





#69
o-Xylene
Concen: 18.157 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion:106 Resp: 58030
Ion Ratio Lower Upper
106 100
91 206.2 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#70
Styrene
Concen: 18.665 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion:104 Resp: 100260
Ion Ratio Lower Upper
104 100
78 45.1 42.4 63.6
103 53.7 44.4 66.6





#71
Bromoform
Concen: 19.238 ug/l
RT: 12.203 min Scan# 1703
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:173 Resp: 22770
Ion Ratio Lower Upper
173 100
175 48.3 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion:152 Resp: 123252
Ion Ratio Lower Upper
152 100
115 53.3 28.2 84.7
150 164.1 0.0 347.6





#73

Isopropylbenzene

Concen: 18.808 ug/l

RT: 12.325 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion:105 Resp: 154099

Ion Ratio Lower Upper

105 100





120 27.0 13.4 40.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#74

N-ethyl acetate

Concen: 19.204 ug/l

RT: 12.142 min Scan# 1693

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Tgt Ion: 43 Resp: 40642




Ion Ratio Lower Upper


43 100

70 43.9 28.7 43.1#

55 27.1 17.3 25.9#

61 23.4 17.0 25.6





#75

1,1,2,2-Tetrachloroethane

Concen: 20.277 ug/l

RT: 12.581 min Scan# 1765

Delta R.T. -0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01





Tgt Ion: 83	Resp: 32678
Ion Ratio	Lower Upper
83 100	
131 10.5	5.6 16.8
85 65.9	32.3 96.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#76

1,2,3-Trichloropropane

Concen: 20.060 ug/l m




RT: 12.630 min Scan# 1773


Delta R.T. 0.001 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31


Tgt Ion: 75	Resp: 22965
Ion Ratio	Lower Upper
75 100	
77 0.0	0.0 0.0





#77
Bromobenzene
Concen: 19.745 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:156 Resp: 4017
Ion Ratio Lower Upper
156 100
77 164.9 95.5 286.4
158 97.5 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#78
n-propylbenzene
Concen: 19.312 ug/l
RT: 12.666 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion: 91 Resp: 188035
Ion Ratio Lower Upper
91 100
120 24.1 11.3 33.9





#79
2-Chlorotoluene
Concen: 18.840 ug/l
RT: 12.752 min Scan# 11
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion: 91 Resp: 103810
Ion Ratio Lower Upper
91 100
126 37.5 16.9 50.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#80
1,3,5-Trimethylbenzene
Concen: 19.194 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion:105 Resp: 130333
Ion Ratio Lower Upper
105 100
120 51.4 24.4 73.4









#81
trans-1,4-Dichloro-2-butene
Concen: 19.407 ug/l
RT: 12.374 min Scan# 1114
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01

Tgt Ion: 75 Resp: 1114
Ion Ratio Lower Upper
75 100
53 87.2 85.3 127.9
89 45.3 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#82
4-Chlorotoluene
Concen: 19.117 ug/l
RT: 12.849 min Scan# 1809
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion: 91 Resp: 109588
Ion Ratio Lower Upper
91 100
126 36.9 17.0 51.0









#83
tert-Butylbenzene
Concen: 18.485 ug/l
RT: 13.075 min Scan# 119
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01

Tgt Ion:119 Resp: 112800
Ion Ratio Lower Upper
119 100
91 59.8 34.8 104.4
134 27.1 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#84
1,2,4-Trimethylbenzene
Concen: 18.868 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31

Tgt Ion:105 Resp: 129138
Ion Ratio Lower Upper
105 100
120 47.2 22.3 66.9





#85

sec-Butylbenzene

Concen: 19.174 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion:105 Resp: 17179

Ion Ratio Lower Upper

105 100





134 21.2 10.2 30.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#86

p-Isopropyltoluene

Concen: 18.794 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31




Tgt Ion:119 Resp: 142418


Ion Ratio Lower Upper

119 100

134 28.1 13.2 39.5


91 22.2 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 19.405 ug/l
RT: 13.361 min Scan# 1893
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:146 Resp: 79720
Ion Ratio Lower Upper
146 100
111 36.2 19.5 58.5
148 63.9 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#88
1,4-Dichlorobenzene
Concen: 19.357 ug/l
RT: 13.440 min Scan# 1906
Delta R.T. -0.006 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Tgt Ion:146 Resp: 80088
Ion Ratio Lower Upper
146 100
111 36.7 19.3 57.8
148 64.3 32.6 97.8





#89
n-Butylbenzene
Concen: 18.890 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01



Tgt Ion: 91 Resp: 13114
Ion Ratio Lower Upper
91 100
92 52.2 26.3 78.9
134 27.1 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#90
Hexachloroethane
Concen: 18.807 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.001 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31


Tgt Ion:117 Resp: 28035
Ion Ratio Lower Upper
117 100
201 83.6 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 19.480 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:146 Resp: 71479
Ion Ratio Lower Upper
146 100
111 37.8 20.3 60.9
148 63.7 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 18.567 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion: 75 Resp: 5028
Ion Ratio Lower Upper
75 100
155 97.9 38.0 113.9
157 125.0 47.8 143.3





#93

1,2,4-Trichlorobenzene

Concen: 18.307 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. 0.000 min

Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Instrument :

MSVOA_Y

ClientSampleId :

VY0729SBS01

Tgt Ion:180 Resp: 42390





Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.3	141.9
145	27.9	14.6	43.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/01/2022

Supervised By :Mahesh Dadoda 08/01/2022



#94

Hexachlorobutadiene

Concen: 18.992 ug/l

RT: 15.111 min Scan# 2180




Delta R.T. -0.000 min


Lab File: VY009768.D

Acq: 29 Jul 2022 13:31

Tgt Ion:225 Resp: 26630


Ion	Ratio	Lower	Upper
225	100		
223	61.5	30.8	92.3
227	64.7	31.6	94.7





#95
Naphthalene
Concen: 17.829 ug/l
RT: 15.233 min Scan# 2200
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31




Instrument :
MSVOA_Y
ClientSampleId :
VY0729SBS01




Tgt Ion:128 Resp: 77390
Ion Ratio Lower Upper
128 100
127 13.3 11.0 16.4
129 11.0 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/01/2022
Supervised By :Mahesh Dadoda 08/01/2022



#96
1,2,3-Trichlorobenzene
Concen: 18.650 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. -0.000 min
Lab File: VY009768.D
Acq: 29 Jul 2022 13:31



Tgt Ion:180 Resp: 37776
Ion Ratio Lower Upper
180 100
182 92.8 47.3 141.9
145 28.1 15.3 45.9

