

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
 Data File : VY009783.D
 Acq On : 29 Jul 2022 19:23
 Operator : KP/MD
 Sample : N3931-04
 Misc : 4.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-116

Quant Time: Aug 01 01:16:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	162347	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	294628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	262866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124995	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	62884	59.138	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.280%
35) Dibromofluoromethane	7.710	113	58614	36.579	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	73.160%
50) Toluene-d8	10.179	98	227723	48.685	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.477	95	65054	30.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	61.240%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072922\
Data File : VY009783.D
Acq On : 29 Jul 2022 19:23
Operator : KP/MD
Sample : N3931-04
Misc : 4.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-116

Quant Time: Aug 01 01:16:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009783.D
Acq: 29 Jul 2022 19:23




Instrument :
MSVOA_Y
ClientSampleId :
RVE-116


Tgt Ion:168 Resp: 162347
Ion Ratio Lower Upper
168 100
99 46.5 46.9 70.3#



#33
1,2-Dichloroethane-d4
Concen: 59.138 ug/l
RT: 8.137 min Scan# 1036
Delta R.T. 0.001 min
Lab File: VY009783.D
Acq: 29 Jul 2022 19:23

Tgt Ion: 65 Resp: 62884
Ion Ratio Lower Upper
65 100
67 55.0 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY009783.D

Acq: 29 Jul 2022 19:23

Instrument :

MSVOA_Y

ClientSampleId :

RVE-116





Tgt Ion:114 Resp: 294628

Ion Ratio Lower Upper

114 100

63 17.0 0.0 43.2

88 13.4 0.0 31.8



#35

Dibromofluoromethane

Concen: 36.579 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.000 min

Lab File: VY009783.D

Acq: 29 Jul 2022 19:23




Tgt Ion:113 Resp: 58614


Ion Ratio Lower Upper

113 100

111 103.4 82.6 124.0

192 22.4 17.6 26.4









#50
Toluene-d8
Concen: 48.685 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009783.D
Acq: 29 Jul 2022 19:23




Instrument :
MSVOA_Y
ClientSampleId :
RVE-116


Tgt Ion: 98 Resp: 227723
Ion Ratio Lower Upper
98 100
100 64.8 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 30.617 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009783.D
Acq: 29 Jul 2022 19:23

Tgt Ion: 95 Resp: 65054
Ion Ratio Lower Upper
95 100
174 99.4 0.0 177.0
176 95.0 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.001 min

Lab File: VY009783.D

Acq: 29 Jul 2022 19:23

Instrument :

MSVOA_Y

ClientSampleId :

RVE-116





Tgt Ion:117 Resp: 262866

Ion Ratio Lower Upper

117 100

82 49.4 43.0 64.4

119 32.7 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY009783.D

Acq: 29 Jul 2022 19:23

Tgt Ion:152 Resp: 124995

Ion Ratio Lower Upper

152 100

115 53.8 28.2 84.7

150 157.4 0.0 347.6

