

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY073120\
 Data File : VY003396.D
 Acq On : 31 Jul 2020 17:57
 Operator : SY/MD
 Sample : L3518-12
 Misc : 3.67G/5ML/MSVOA Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-6

Quant Time: Aug 01 02:27:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y073120S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 31 13:09:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	403235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	628694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	565669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	261251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.14	65	210311	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
35) Dibromofluoromethane	7.72	113	198546	53.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.22%	
50) Toluene-d8	10.18	98	764511	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
62) 4-Bromofluorobenzene	12.48	95	238481	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
Target Compounds						
16) Acetone	3.96	43	223269	235.690	ug/l	96
20) Methylene Chloride	4.72	84	68293	6.632	ug/l	96
25) 2-Butanone	7.00	43	71018	49.195	ug/l	96
52) Toluene	10.24	92	63040	6.089	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY073120\
Data File : VY003396.D
Acq On : 31 Jul 2020 17:57
Operator : SY/MD
Sample : L3518-12
Misc : 3.67G/5ML/MSVOA Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-6

Quant Time: Aug 01 02:27:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y073120S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 31 13:09:26 2020
Response via : Initial Calibration

