

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
 Data File : VY009790.D
 Acq On : 01 Aug 2022 12:35
 Operator : KP/MD
 Sample : VY0801SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0801SBL01

Quant Time: Aug 02 01:36:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	176456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	283485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	264818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124789	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	75870	66.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.560%
35) Dibromofluoromethane	7.710	113	74447	48.286	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.580%
50) Toluene-d8	10.179	98	244326	54.821	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	109.640%
62) 4-Bromofluorobenzene	12.483	95	74383	36.384	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	72.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
Data File : VY009790.D
Acq On : 01 Aug 2022 12:35
Operator : KP/MD
Sample : VY0801SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0801SBL01

Quant Time: Aug 02 01:36:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

