

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
 Data File : VY009801.D
 Acq On : 01 Aug 2022 17:02
 Operator : KP/MD
 Sample : N3976-01
 Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP1-0729

Quant Time: Aug 02 01:40:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	134909	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	235322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	255993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	119881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	59033	67.554	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.100%
35) Dibromofluoromethane	7.710	113	58220	45.490	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.980%
50) Toluene-d8	10.179	98	235825	64.498	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	129.000%#
62) 4-Bromofluorobenzene	12.477	95	76269	44.942	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	89.880%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.704	84	37984	18.413	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
Data File : VY009801.D
Acq On : 01 Aug 2022 17:02
Operator : KP/MD
Sample : N3976-01
Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP1-0729

Quant Time: Aug 02 01:40:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

