

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
 Data File : VY009806.D
 Acq On : 01 Aug 2022 18:59
 Operator : KP/MD
 Sample : N3960-04
 Misc : 2.83g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-152

Quant Time: Aug 02 01:42:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	132939	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	236353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	267615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	112276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	60491	70.477	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.960%
35) Dibromofluoromethane	7.710	113	57436	44.681	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.360%
50) Toluene-d8	10.179	98	239161	65.170	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	130.340%#
62) 4-Bromofluorobenzene	12.483	95	64923	38.089	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	76.180%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.710	84	24533	7.931	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080122\
Data File : VY009806.D
Acq On : 01 Aug 2022 18:59
Operator : KP/MD
Sample : N3960-04
Misc : 2.83g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-152

Quant Time: Aug 02 01:42:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

