

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080221\
 Data File : VY005513.D
 Acq On : 02 Aug 2021 09:24
 Operator : SY/MD
 Sample : VSTDIC005
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDIC005

Quant Time: Aug 03 04:34:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:33:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	5.350	89.3#	12	0.00
3 P	Chloromethane	50.000	5.233	89.5#	11	0.00
4 C	Vinyl Chloride	50.000	5.335	89.3#	11	0.00
5 T	Bromomethane	50.000	5.625	88.8#	12	0.00
6 T	Chloroethane	50.000	5.200	89.6#	10	0.00
7 T	Trichlorofluoromethane	50.000	5.287	89.4#	10	0.00
8 T	Diethyl Ether	50.000	5.386	89.2#	12	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	5.324	89.4#	11	0.00
10 T	Methyl Iodide	50.000	4.986	90.0#	10	0.00
11 T	Tert butyl alcohol	250.000	30.524	87.8#	15	0.01
12 CM	1,1-Dichloroethene	50.000	5.391	89.2#	11	0.00
13 T	Acrolein	250.000	23.027	90.8#	9	0.00
14 T	Allyl chloride	50.000	5.047	89.9#	10	0.00
15 T	Acrylonitrile	250.000	25.580	89.8#	12	0.00
16 T	Acetone	250.000	25.881	89.6#	11	0.00
17 T	Carbon Disulfide	50.000	5.356	89.3#	11	0.00
18 T	Methyl Acetate	50.000	6.126	87.7#	15	0.00
19 T	Methyl tert-butyl Ether	50.000	5.116	89.8#	11	0.00
20 T	Methylene Chloride	50.000	8.376	83.2#	21	0.00
21 T	trans-1,2-Dichloroethene	50.000	5.366	89.3#	11	0.00
22 T	Diisopropyl ether	50.000	5.024	90.0#	10	0.00
23 T	Vinyl Acetate	250.000	24.002	90.4#	10	0.00
24 P	1,1-Dichloroethane	50.000	5.085	89.8#	10	0.00
25 T	2-Butanone	250.000	25.121	90.0#	11	0.00
26 T	2,2-Dichloropropane	50.000	5.259	89.5#	11	0.00
27 T	cis-1,2-Dichloroethene	50.000	5.237	89.5#	11	0.00
28 T	Bromochloromethane	50.000	5.313	89.4#	12	0.00
29 T	Tetrahydrofuran	250.000	25.987	89.6#	12	0.00
30 C	Chloroform	50.000	5.310	89.4#	11	0.00
31 T	Cyclohexane	50.000	5.811	88.4#	13	0.00
32 T	1,1,1-Trichloroethane	50.000	5.141	89.7#	11	0.00
33 S	1,2-Dichloroethane-d4	50.000	5.276	89.4#	11	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	5.028	89.9#	10	0.00
36 T	1,1-Dichloropropene	50.000	5.119	89.8#	11	0.00
37 T	Ethyl Acetate	50.000	5.295	89.4#	12	0.00
38 T	Carbon Tetrachloride	50.000	5.190	89.6#	11	0.00
39 T	Methylcyclohexane	50.000	4.965	90.1#	10	0.00
40 TM	Benzene	50.000	5.208	89.6#	11	0.00
41 T	Methacrylonitrile	50.000	3.673	92.7#	8	-0.01
42 TM	1,2-Dichloroethane	50.000	5.248	89.5#	11	0.00
43 T	Isopropyl Acetate	50.000	5.023	90.0#	12	0.00
44 TM	Trichloroethene	50.000	5.192	89.6#	11	0.00
45 C	1,2-Dichloropropane	50.000	5.190	89.6#	11	0.00
46 T	Dibromomethane	50.000	5.129	89.7#	11	0.00
47 T	Bromodichloromethane	50.000	5.055	89.9#	11	0.00
48 T	Methyl methacrylate	50.000	4.945	90.1#	11	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080221\
 Data File : VY005513.D
 Acq On : 02 Aug 2021 09:24
 Operator : SY/MD
 Sample : VSTDIC005
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDIC005

Quant Time: Aug 03 04:34:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:33:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	102.570	89.7#	12	0.00
50 S	Toluene-d8	50.000	5.152	89.7#	10	0.00
51 T	4-Methyl-2-Pentanone	250.000	25.044	90.0#	12	0.00
52 CM	Toluene	50.000	5.124	89.8#	11	0.00
53 T	t-1,3-Dichloropropene	50.000	5.098	89.8#	11	0.00
54 T	cis-1,3-Dichloropropene	50.000	5.024	90.0#	10	0.00
55 T	1,1,2-Trichloroethane	50.000	5.103	89.8#	11	0.00
56 T	Ethyl methacrylate	50.000	4.871	90.3#	11	0.00
57 T	1,3-Dichloropropane	50.000	5.190	89.6#	11	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	23.988	90.4#	10	0.00
59 T	2-Hexanone	250.000	24.253	90.3#	11	0.00
60 T	Dibromochloromethane	50.000	5.093	89.8#	11	0.00
61 T	1,2-Dibromoethane	50.000	5.095	89.8#	11	0.00
62 S	4-Bromofluorobenzene	50.000	5.398	89.2#	11	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	103	0.00
64 T	Tetrachloroethene	50.000	5.444	89.1#	11	0.00
65 PM	Chlorobenzene	50.000	5.139	89.7#	11	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	5.069	89.9#	11	0.00
67 C	Ethyl Benzene	50.000	5.017	90.0#	10	0.00
68 T	m/p-Xylenes	100.000	9.942	90.1#	10	0.00
69 T	o-Xylene	50.000	5.088	89.8#	10	0.00
70 T	Styrene	50.000	4.961	90.1#	10	0.00
71 P	Bromoform	50.000	5.054	89.9#	11	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	4.899	90.2#	10	0.00
74 T	N-ethyl acetate	50.000	4.710	90.6#	10	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	5.048	89.9#	11	0.00
76 T	1,2,3-Trichloropropane	50.000	6.287	87.4#	12	0.00
77 T	Bromobenzene	50.000	4.997	90.0#	10	0.00
78 T	n-propylbenzene	50.000	4.948	90.1#	10	0.00
79 T	2-Chlorotoluene	50.000	5.002	90.0#	10	0.00
80 T	1,3,5-Trimethylbenzene	50.000	4.953	90.1#	10	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	5.108	89.8#	12	0.00
82 T	4-Chlorotoluene	50.000	5.020	90.0#	10	0.00
83 T	tert-Butylbenzene	50.000	4.982	90.0#	10	0.00
84 T	1,2,4-Trimethylbenzene	50.000	4.870	90.3#	10	0.00
85 T	sec-Butylbenzene	50.000	4.969	90.1#	10	0.00
86 T	p-Isopropyltoluene	50.000	4.911	90.2#	10	0.00
87 T	1,3-Dichlorobenzene	50.000	5.054	89.9#	11	0.00
88 T	1,4-Dichlorobenzene	50.000	5.188	89.6#	11	0.00
89 T	n-Butylbenzene	50.000	4.943	90.1#	10	0.00
90 T	Hexachloroethane	50.000	5.025	90.0#	10	0.00
91 T	1,2-Dichlorobenzene	50.000	5.213	89.6#	11	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	5.070	89.9#	12	0.00
93 T	1,2,4-Trichlorobenzene	50.000	4.989	90.0#	11	0.00
94 T	Hexachlorobutadiene	50.000	5.197	89.6#	11	0.00
95 T	Naphthalene	50.000	4.950	90.1#	11	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080221\
Data File : VY005513.D
Acq On : 02 Aug 2021 09:24
Operator : SY/MD
Sample : VSTDIC005
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDIC005

Quant Time: Aug 03 04:34:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:33:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	4.983	90.0#	11	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6