

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
 Data File : VY014932.D
 Acq On : 02 Aug 2023 15:42
 Operator : SY/MD
 Sample : 03862-01
 Misc : 7.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BUILDING-B-11-(10-15)

Quant Time: Aug 03 02:43:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	212094	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	369461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	340554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	158282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	130570	54.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.040%	
35) Dibromofluoromethane	7.710	113	114073	48.502	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.000%	
50) Toluene-d8	10.179	98	434217	50.077	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.160%	
62) 4-Bromofluorobenzene	12.477	95	151623	45.604	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.200%	
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	30175	8.634	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
Data File : VY014932.D
Acq On : 02 Aug 2023 15:42
Operator : SY/MD
Sample : 03862-01
Misc : 7.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BUILDING-B-11-(10-15)

Quant Time: Aug 03 02:43:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

