

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
 Data File : VY014935.D
 Acq On : 02 Aug 2023 16:53
 Operator : SY/MD
 Sample : 03862-07
 Misc : 5.78g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BUILDING-B-15-(5-10)

Quant Time: Aug 03 02:45:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	219533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	386299	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	365665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	161716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	144897	57.920	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.840%
35) Dibromofluoromethane	7.710	113	122010	49.616	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.240%
50) Toluene-d8	10.173	98	459281	50.659	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.320%
62) 4-Bromofluorobenzene	12.477	95	167478	48.177	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.360%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.192	76	12946	2.092	ug/l	99
20) Methylene Chloride	4.710	84	29950	8.090	ug/l	92
30) Chloroform	7.502	83	31405	6.820	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
Data File : VY014935.D
Acq On : 02 Aug 2023 16:53
Operator : SY/MD
Sample : 03862-07
Misc : 5.78g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BUILDING-B-15-(5-10)

Quant Time: Aug 03 02:45:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

