

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
 Data File : VY014928.D
 Acq On : 02 Aug 2023 11:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:41:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.407	0.350	14.0	74	0.00
3 P	Chloromethane	0.458	0.489	-6.8	90	0.00
4 C	Vinyl Chloride	0.512	0.588	-14.8#	94	0.00
5 T	Bromomethane	0.338	0.415	-22.8	102	0.00
6 T	Chloroethane	0.320	0.401	-25.3#	101	0.00
7 T	Trichlorofluoromethane	0.860	0.897	-4.3	84	0.00
8 T	Diethyl Ether	0.320	0.317	0.9	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.534	-1.3	81	0.01
10 T	Methyl Iodide	0.493	0.609	-23.5	92	0.00
11 T	Tert butyl alcohol	0.125	0.082	34.4#	61	0.01
12 CM	1,1-Dichloroethene	0.492	0.480	2.4#	78	0.00
13 T	Acrolein	0.060	0.032	46.7#	49#	0.00
14 T	Allyl chloride	0.874	0.792	9.4	74	0.00
15 T	Acrylonitrile	0.176	0.179	-1.7	82	0.00
16 T	Acetone	0.181	0.189	-4.4	80	0.00
17 T	Carbon Disulfide	1.410	1.180	16.3	70	0.00
18 T	Methyl Acetate	0.693	0.701	-1.2	82	0.00
19 T	Methyl tert-butyl Ether	1.599	1.610	-0.7	81	0.00
20 T	Methylene Chloride	0.656	0.661	-0.8	90	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.549	1.8	80	0.00
22 T	Diisopropyl ether	1.777	1.762	0.8	79	0.00
23 T	Vinyl Acetate	1.064	1.030	3.2	77	0.00
24 P	1,1-Dichloroethane	1.017	1.029	-1.2	81	0.00
25 T	2-Butanone	0.264	0.263	0.4	80	0.00
26 T	2,2-Dichloropropane	0.969	0.974	-0.5	81	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.670	-3.1	83	0.00
28 T	Bromochloromethane	0.438	0.422	3.7	83	0.00
29 T	Tetrahydrofuran	0.163	0.156	4.3	80	0.00
30 C	Chloroform	1.049	1.085	-3.4#	84	0.00
31 T	Cyclohexane	0.946	0.818	13.5	74	0.00
32 T	1,1,1-Trichloroethane	0.944	0.973	-3.1	83	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.552	3.2	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.318	0.309	2.8	79	0.00
36 T	1,1-Dichloropropene	0.476	0.477	-0.2	80	0.00
37 T	Ethyl Acetate	0.328	0.320	2.4	82	0.00
38 T	Carbon Tetrachloride	0.503	0.532	-5.8	84	0.00
39 T	Methylcyclohexane	0.620	0.585	5.6	75	0.00
40 TM	Benzene	1.354	1.376	-1.6	81	0.00
41 T	Methacrylonitrile	0.169	0.185	-9.5	91	0.00
42 TM	1,2-Dichloroethane	0.429	0.448	-4.4	83	0.00
43 T	Isopropyl Acetate	0.624	0.606	2.9	78	0.00
44 TM	Trichloroethene	0.379	0.398	-5.0	84	0.00
45 C	1,2-Dichloropropane	0.353	0.358	-1.4#	80	0.00
46 T	Dibromomethane	0.214	0.219	-2.3	81	0.00
47 T	Bromodichloromethane	0.505	0.525	-4.0	82	0.00
48 T	Methyl methacrylate	0.281	0.269	4.3	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
 Data File : VY014928.D
 Acq On : 02 Aug 2023 11:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:41:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	79	0.00
50 S	Toluene-d8	1.173	1.090	7.1	77	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.332	-1.2	81	0.00
52 CM	Toluene	0.866	0.881	-1.7#	81	0.00
53 T	t-1,3-Dichloropropene	0.553	0.545	1.4	78	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.599	0.3	79	0.00
55 T	1,1,2-Trichloroethane	0.299	0.311	-4.0	83	0.00
56 T	Ethyl methacrylate	0.456	0.452	0.9	78	0.00
57 T	1,3-Dichloropropane	0.515	0.528	-2.5	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.201	5.6	77	0.00
59 T	2-Hexanone	0.239	0.242	-1.3	80	0.00
60 T	Dibromochloromethane	0.366	0.385	-5.2	83	0.00
61 T	1,2-Dibromoethane	0.298	0.304	-2.0	81	0.00
62 S	4-Bromofluorobenzene	0.450	0.404	10.2	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.402	0.503	-25.1#	101	0.00
65 PM	Chlorobenzene	1.065	1.094	-2.7	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.421	-4.7	83	0.00
67 C	Ethyl Benzene	1.924	1.953	-1.5#	81	0.00
68 T	m/p-Xylenes	0.734	0.750	-2.2	82	0.00
69 T	o-Xylene	0.715	0.729	-2.0	82	0.00
70 T	Styrene	1.206	1.236	-2.5	81	0.00
71 P	Bromoform	0.272	0.278	-2.2	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.915	3.906	0.2	83	0.00
74 T	N-amyl acetate	1.311	1.210	7.7	76	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.808	9.5	76	0.00
76 T	1,2,3-Trichloropropane	0.631	0.619	1.9	82	0.00
77 T	Bromobenzene	0.894	0.912	-2.0	84	0.00
78 T	n-propylbenzene	4.718	4.642	1.6	81	0.00
79 T	2-Chlorotoluene	2.717	2.612	3.9	79	0.00
80 T	1,3,5-Trimethylbenzene	3.270	3.224	1.4	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.284	12.9	71	0.00
82 T	4-Chlorotoluene	2.799	2.688	4.0	79	0.00
83 T	tert-Butylbenzene	2.900	2.871	1.0	81	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.170	1.2	81	0.00
85 T	sec-Butylbenzene	4.255	4.287	-0.8	83	0.00
86 T	p-Isopropyltoluene	3.474	3.526	-1.5	83	0.00
87 T	1,3-Dichlorobenzene	1.739	1.786	-2.7	84	0.00
88 T	1,4-Dichlorobenzene	1.743	1.764	-1.2	83	0.00
89 T	n-Butylbenzene	3.380	3.362	0.5	81	0.00
90 T	Hexachloroethane	0.714	0.691	3.2	79	0.00
91 T	1,2-Dichlorobenzene	1.594	1.629	-2.2	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.156	8.8	77	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.026	-1.4	81	0.00
94 T	Hexachlorobutadiene	0.556	0.568	-2.2	82	0.00
95 T	Naphthalene	2.357	2.351	0.3	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
Data File : VY014928.D
Acq On : 02 Aug 2023 11:32
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 03 02:41:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.909	-1.1	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6