

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
 Data File : VY014929.D
 Acq On : 02 Aug 2023 14:15
 Operator : SY/MD
 Sample : VY0802SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0802SBL01

Quant Time: Aug 03 02:42:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	286229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	482331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	417221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	202428	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	169214	51.879	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.760%
35) Dibromofluoromethane	7.716	113	153713	50.063	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.120%
50) Toluene-d8	10.179	98	537802	47.509	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.020%
62) 4-Bromofluorobenzene	12.477	95	192527	44.356	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.720%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080223\
Data File : VY014929.D
Acq On : 02 Aug 2023 14:15
Operator : SY/MD
Sample : VY0802SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0802SBL01

Quant Time: Aug 03 02:42:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY014929.D
Acq: 02 Aug 2023 14:15




Instrument :
MSVOA_Y
ClientSampleId :
VY0802SBL01


Tgt Ion:168 Resp: 286229
Ion Ratio Lower Upper
168 100
99 56.9 43.4 65.0



#33
1,2-Dichloroethane-d4
Concen: 51.879 ug/l
RT: 8.137 min Scan# 1036
Delta R.T. 0.001 min
Lab File: VY014929.D
Acq: 02 Aug 2023 14:15

Tgt Ion: 65 Resp: 169214
Ion Ratio Lower Upper
65 100
67 51.9 0.0 101.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY014929.D

Acq: 02 Aug 2023 14:15

Instrument :

MSVOA_Y

ClientSampleId :

VY0802SBL01





Tgt Ion:114 Resp: 482331

Ion Ratio Lower Upper

114 100

63 19.4 0.0 42.4

88 15.5 0.0 30.6



#35

Dibromofluoromethane

Concen: 50.063 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY014929.D

Acq: 02 Aug 2023 14:15




Tgt Ion:113 Resp: 153713


Ion Ratio Lower Upper

113 100

111 101.4 81.3 121.9

192 18.9 16.9 25.3









#50
Toluene-d8
Concen: 47.509 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY014929.D
Acq: 02 Aug 2023 14:15




Instrument :
MSVOA_Y
ClientSampleId :
VY0802SBL01


Tgt Ion: 98 Resp: 537802
Ion Ratio Lower Upper
98 100
100 65.5 50.1 75.1



#62
4-Bromofluorobenzene
Concen: 44.356 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY014929.D
Acq: 02 Aug 2023 14:15

Tgt Ion: 95 Resp: 192527
Ion Ratio Lower Upper
95 100
174 81.3 0.0 178.2
176 79.1 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.001 min

Lab File: VY014929.D

Acq: 02 Aug 2023 14:15

Instrument :

MSVOA_Y

ClientSampleId :

VY0802SBL01





Tgt Ion:117 Resp: 417221

Ion Ratio Lower Upper

117 100

82 55.6 43.4 65.0

119 33.5 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY014929.D

Acq: 02 Aug 2023 14:15

Tgt Ion:152 Resp: 202428

Ion Ratio Lower Upper

152 100

115 58.7 28.8 86.5

150 156.4 0.0 348.8

