

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY080320\
 Data File : VY003416.D
 Acq On : 03 Aug 2020 14:51
 Operator : SY/MD
 Sample : L3521-03
 Misc : 8.43G/5ML/MSVOA Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-AB

Quant Time: Aug 04 02:27:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y073120S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 31 13:09:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	453532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	706539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	658285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	306022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.14	65	226457	52.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.02%	
35) Dibromofluoromethane	7.72	113	224489	53.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.88%	
50) Toluene-d8	10.18	98	872936	53.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.12%	
62) 4-Bromofluorobenzene	12.48	95	283424	50.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.80%	
Target Compounds						
3) Chloromethane	2.12	50	11357	2.233	ug/l	95
16) Acetone	3.96	43	40176	37.708	ug/l	98
25) 2-Butanone	6.99	43	13017	8.017	ug/l	95
40) Benzene	8.16	78	21551	1.162	ug/l	94
52) Toluene	10.24	92	24451	2.102	ug/l	99
68) m/p-Xylenes	11.69	106	22447	2.556	ug/l	99
69) o-Xylene	12.02	106	10597	1.289	ug/l	99
84) 1,2,4-Trimethylbenzene	13.12	105	23360	1.400	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY080320\
Data File : VY003416.D
Acq On : 03 Aug 2020 14:51
Operator : SY/MD
Sample : L3521-03
Misc : 8.43G/5ML/MSVOA Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-AB

Quant Time: Aug 04 02:27:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y073120S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 31 13:09:26 2020
Response via : Initial Calibration

