


# Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 4.253\text{e-}001 * \text{Amt} + 3.679\text{e-}002$$


Coef of Det (r^2) = 0.999661 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

# Naphthalene

Response Ratio



$$\text{Response} = 2.227\text{e}+000 * \text{Amt} - 2.414\text{e}-001$$


Coef of Det ( $r^2$ ) = 0.994551 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

# N-amyl acetate

Response Ratio



$$\text{Response} = 1.335\text{e}+000 * \text{Amt} - 9.191\text{e}-002$$


Coef of Det ( $r^2$ ) = 0.999587 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

# Styrene

Response Ratio



$$\text{Response} = 1.262\text{e}+000 * \text{Amt} - 2.452\text{e}-002$$


Coef of Det ( $r^2$ ) = 0.997981 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

## Ethyl methacrylate

Response Ratio



$$\text{Response} = 4.719\text{e-}001 * \text{Amt} - 2.780\text{e-}002$$


Coef of Det ( $r^2$ ) = 0.997265 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

## Toluene-d8

Response Ratio




$$\text{Response} = 7.712\text{e-}001 * \text{Amt} + 1.237\text{e-}001$$

Coef of Det (r^2) = 0.991626 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022


## 1,4-Dioxane



Response =  $3.200\text{e-}003 \times \text{Amt} - 4.021\text{e-}003$   
Coef of Det ( $r^2$ ) = 0.995922 Curve Fit: Linear  
Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M  
Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

## 1,2-Dichloroethane-d4

Response Ratio



$$\text{Response} = 3.611\text{e-}001 * \text{Amt} + 4.993\text{e-}002$$


Coef of Det (r^2) = 0.996103 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

# Methylene Chloride

Response Ratio



$$\text{Response} = 5.478\text{e-}001 * \text{Amt} + 1.038\text{e-}001$$


Coef of Det ( $r^2$ ) = 0.999579 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M

Calibration Table Last Updated: Thu Aug 04 02:27:45 2022

# Bromomethane

Response Ratio



$R = -6.073e-002 A^2 + 5.635e-001 A + 2.299e-003$   
 Coef of Det ( $r^2$ ) = 0.999477    Curve Fit: Quadratic  
 Method Name:    Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y080322S.M  
 Calibration Table Last Updated: Thu Aug 04 02:27:45 2022