



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080322\
Data File : VY009813.D
Acq On : 03 Aug 2022 11:06
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Title : SW846 8260
Last Update : Thu Aug 04 02:27:45 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	1734	PASS
75	95	30	60	53.2	4397	PASS
95	95	100	100	100.0	8269	PASS
96	95	5	9	7.4	616	PASS
173	174	0.00	2	1.1	90	PASS
174	95	50	100	97.1	8027	PASS
175	174	5	9	8.2	661	PASS
176	174	95	101	97.6	7832	PASS
177	176	5	9	6.3	496	PASS