

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
 Data File : VY014951.D
 Acq On : 03 Aug 2023 16:15
 Operator : SY/MD
 Sample : 03877-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1RE

Quant Time: Aug 04 04:32:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	87104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	150468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	141402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	66119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	57707	58.138	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.280%
35) Dibromofluoromethane	7.716	113	47804	49.908	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.820%
50) Toluene-d8	10.179	98	178298	50.490	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.980%
62) 4-Bromofluorobenzene	12.477	95	63984	47.253	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.500%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.722	84	11039	7.188	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
Data File : VY014951.D
Acq On : 03 Aug 2023 16:15
Operator : SY/MD
Sample : 03877-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1RE

Quant Time: Aug 04 04:32:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

