

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
 Data File : VY014959.D
 Acq On : 03 Aug 2023 19:23
 Operator : SY/MD
 Sample : 03897-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Aug 04 04:34:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	76200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	129943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	110726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	40902	47.104	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.200%
35) Dibromofluoromethane	7.716	113	40354	48.785	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.560%
50) Toluene-d8	10.179	98	148713	48.764	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.483	95	46675	39.915	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
Data File : VY014959.D
Acq On : 03 Aug 2023 19:23
Operator : SY/MD
Sample : 03897-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Aug 04 04:34:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

