

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
 Data File : VY014945.D
 Acq On : 03 Aug 2023 11:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 04 04:30:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.407	0.334	17.9	72	0.00
3 P	Chloromethane	0.458	0.452	1.3	84	0.00
4 C	Vinyl Chloride	0.512	0.566	-10.5#	91	0.00
5 T	Bromomethane	0.338	0.427	-26.3#	106	0.00
6 T	Chloroethane	0.320	0.399	-24.7	102	0.00
7 T	Trichlorofluoromethane	0.860	0.891	-3.6	85	0.00
8 T	Diethyl Ether	0.320	0.326	-1.9	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.538	-2.1	83	0.00
10 T	Methyl Iodide	0.493	0.637	-29.2#	98	0.00
11 T	Tert butyl alcohol	0.125	0.118	5.6	89	0.00
12 CM	1,1-Dichloroethene	0.492	0.490	0.4#	81	0.00
13 T	Acrolein	0.060	0.032	46.7#	49#	0.00
14 T	Allyl chloride	0.874	0.801	8.4	75	0.01
15 T	Acrylonitrile	0.176	0.199	-13.1	93	0.00
16 T	Acetone	0.181	0.213	-17.7	92	0.00
17 T	Carbon Disulfide	1.410	1.143	18.9	68	0.00
18 T	Methyl Acetate	0.693	0.784	-13.1	93	0.00
19 T	Methyl tert-butyl Ether	1.599	1.735	-8.5	88	0.00
20 T	Methylene Chloride	0.656	0.632	3.7	87	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.558	0.2	82	0.00
22 T	Diisopropyl ether	1.777	1.796	-1.1	82	0.00
23 T	Vinyl Acetate	1.064	1.079	-1.4	82	0.00
24 P	1,1-Dichloroethane	1.017	1.056	-3.8	84	0.00
25 T	2-Butanone	0.264	0.298	-12.9	92	0.00
26 T	2,2-Dichloropropane	0.969	1.016	-4.9	86	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.699	-7.5	87	0.00
28 T	Bromochloromethane	0.438	0.455	-3.9	91	0.00
29 T	Tetrahydrofuran	0.163	0.174	-6.7	90	0.00
30 C	Chloroform	1.049	1.146	-9.2#	89	0.00
31 T	Cyclohexane	0.946	0.782	17.3	71	0.00
32 T	1,1,1-Trichloroethane	0.944	1.004	-6.4	87	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.606	-6.3	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.318	0.342	-7.5	90	0.00
36 T	1,1-Dichloropropene	0.476	0.466	2.1	80	0.00
37 T	Ethyl Acetate	0.328	0.349	-6.4	91	0.00
38 T	Carbon Tetrachloride	0.503	0.531	-5.6	86	0.00
39 T	Methylcyclohexane	0.620	0.568	8.4	75	0.00
40 TM	Benzene	1.354	1.388	-2.5	84	0.00
41 T	Methacrylonitrile	0.169	0.157	7.1	79	0.00
42 TM	1,2-Dichloroethane	0.429	0.465	-8.4	88	0.00
43 T	Isopropyl Acetate	0.624	0.649	-4.0	86	0.00
44 TM	Trichloroethene	0.379	0.402	-6.1	87	0.00
45 C	1,2-Dichloropropane	0.353	0.371	-5.1#	85	0.00
46 T	Dibromomethane	0.214	0.231	-7.9	88	0.00
47 T	Bromodichloromethane	0.505	0.544	-7.7	87	0.00
48 T	Methyl methacrylate	0.281	0.284	-1.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
 Data File : VY014945.D
 Acq On : 03 Aug 2023 11:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 04 04:30:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	86	0.00
50 S	Toluene-d8	1.173	1.169	0.3	84	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.363	-10.7	91	0.00
52 CM	Toluene	0.866	0.901	-4.0#	85	0.00
53 T	t-1,3-Dichloropropene	0.553	0.569	-2.9	84	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.618	-2.8	84	0.00
55 T	1,1,2-Trichloroethane	0.299	0.331	-10.7	90	0.00
56 T	Ethyl methacrylate	0.456	0.478	-4.8	84	0.00
57 T	1,3-Dichloropropane	0.515	0.548	-6.4	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.211	0.9	83	0.00
59 T	2-Hexanone	0.239	0.269	-12.6	92	0.00
60 T	Dibromochloromethane	0.366	0.404	-10.4	89	0.00
61 T	1,2-Dibromoethane	0.298	0.323	-8.4	88	0.00
62 S	4-Bromofluorobenzene	0.450	0.436	3.1	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.402	0.442	-10.0	92	0.00
65 PM	Chlorobenzene	1.065	1.133	-6.4	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.437	-8.7	89	0.00
67 C	Ethyl Benzene	1.924	1.989	-3.4#	85	0.00
68 T	m/p-Xylenes	0.734	0.756	-3.0	85	0.00
69 T	o-Xylene	0.715	0.751	-5.0	86	0.00
70 T	Styrene	1.206	1.275	-5.7	86	0.00
71 P	Bromoform	0.272	0.305	-12.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.915	3.912	0.1	87	0.00
74 T	N-amyl acetate	1.311	1.263	3.7	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.917	-2.7	90	0.00
76 T	1,2,3-Trichloropropane	0.631	0.679	-7.6	94	0.00
77 T	Bromobenzene	0.894	0.927	-3.7	89	0.00
78 T	n-propylbenzene	4.718	4.605	2.4	83	0.00
79 T	2-Chlorotoluene	2.717	2.652	2.4	84	0.00
80 T	1,3,5-Trimethylbenzene	3.270	3.255	0.5	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.299	8.3	78	0.00
82 T	4-Chlorotoluene	2.799	2.721	2.8	83	0.00
83 T	tert-Butylbenzene	2.900	2.887	0.4	85	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.216	-0.2	86	0.00
85 T	sec-Butylbenzene	4.255	4.226	0.7	85	0.00
86 T	p-Isopropyltoluene	3.474	3.537	-1.8	86	0.00
87 T	1,3-Dichlorobenzene	1.739	1.843	-6.0	90	0.00
88 T	1,4-Dichlorobenzene	1.743	1.820	-4.4	89	0.00
89 T	n-Butylbenzene	3.380	3.275	3.1	82	0.00
90 T	Hexachloroethane	0.714	0.690	3.4	82	0.00
91 T	1,2-Dichlorobenzene	1.594	1.682	-5.5	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.174	-1.8	89	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.064	-5.1	87	0.00
94 T	Hexachlorobutadiene	0.556	0.551	0.9	82	0.00
95 T	Naphthalene	2.357	2.531	-7.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
Data File : VY014945.D
Acq On : 03 Aug 2023 11:27
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 04 04:30:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.944	-5.0	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6