



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080323\
Data File : VY014944.D
Acq On : 03 Aug 2023 08:59
Operator : SY/MD
Sample : BFB
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Title : SW846 8260
Last Update : Fri Jul 14 02:09:50 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.6	23419	PASS
75	95	30	60	53.6	63949	PASS
95	95	100	100	100.0	119360	PASS
96	95	5	9	6.6	7836	PASS
173	174	0.00	2	1.0	954	PASS
174	95	50	100	80.9	96576	PASS
175	174	5	9	7.6	7372	PASS
176	174	95	101	97.2	93859	PASS
177	176	5	9	6.9	6431	PASS