

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080421\
 Data File : VY005540.D
 Acq On : 04 Aug 2021 16:11
 Operator : SY/MD
 Sample : M3291-04
 Misc : 6.20G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2

Quant Time: Aug 05 06:40:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	17531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	28663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	29287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10497	59.152	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.300%
35) Dibromofluoromethane	7.728	113	7804	48.630	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.260%
50) Toluene-d8	10.185	98	37461	55.660	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.320%
62) 4-Bromofluorobenzene	12.483	95	14096	63.247	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	126.500%
Target Compounds						
					Qvalue	
10) Methyl Iodide	4.101	142	1139	6.111	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080421\
Data File : VY005540.D
Acq On : 04 Aug 2021 16:11
Operator : SY/MD
Sample : M3291-04
Misc : 6.20G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2

Quant Time: Aug 05 06:40:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

