

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080422\
 Data File : VY009835.D
 Acq On : 04 Aug 2022 16:21
 Operator : KP/MD
 Sample : N3961-14
 Misc : 3.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-180

Quant Time: Aug 05 01:44:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 04 02:27:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	47797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	80308	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	77077	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	36949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	24304	63.493	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.980%
35) Dibromofluoromethane	7.710	113	22465	45.880	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.760%
50) Toluene-d8	10.179	98	69220	47.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	95.720%
62) 4-Bromofluorobenzene	12.477	95	21339	32.183	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	64.360%
Target Compounds						
20) Methylene Chloride	4.692	84	8111	6.011	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080422\
Data File : VY009835.D
Acq On : 04 Aug 2022 16:21
Operator : KP/MD
Sample : N3961-14
Misc : 3.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-180

Quant Time: Aug 05 01:44:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 04 02:27:45 2022
Response via : Initial Calibration

