

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
 Data File : VY005572.D
 Acq On : 06 Aug 2021 15:49
 Operator : SY/MD
 Sample : M3307-01
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ219-367

Quant Time: Aug 07 06:19:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	265193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	448872	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	456806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	234244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	155694	58.00	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.00%
35) Dibromofluoromethane	7.728	113	134763	53.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.24%
50) Toluene-d8	10.185	98	582306	55.25	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.50%
62) 4-Bromofluorobenzene	12.483	95	212311	60.83	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.66%
Target Compounds						
20) Methylene Chloride	4.728	84	41480	10.69	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
Data File : VY005572.D
Acq On : 06 Aug 2021 15:49
Operator : SY/MD
Sample : M3307-01
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ219-367

Quant Time: Aug 07 06:19:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

