

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
 Data File : VY005575.D
 Acq On : 06 Aug 2021 16:59
 Operator : SY/MD
 Sample : M3308-01
 Misc : 5.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-325

Quant Time: Aug 07 06:20:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	252592	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	428037	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	383935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	145346	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	147793	57.80	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.60%
35) Dibromofluoromethane	7.728	113	126958	52.98	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.96%
50) Toluene-d8	10.184	98	520184	51.76	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.52%
62) 4-Bromofluorobenzene	12.483	95	160539	48.23	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.46%
Target Compounds						
20) Methylene Chloride	4.728	84	45369	13.30	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
Data File : VY005575.D
Acq On : 06 Aug 2021 16:59
Operator : SY/MD
Sample : M3308-01
Misc : 5.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-325

Quant Time: Aug 07 06:20:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

