

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
 Data File : VY005576.D
 Acq On : 06 Aug 2021 17:21
 Operator : SY/MD
 Sample : M3308-04
 Misc : 1.19G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11930

Quant Time: Aug 07 06:20:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	54097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	107888	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	52656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	15968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	73512	134.24	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 268.48%#			
35) Dibromofluoromethane	7.728	113	48382	80.10	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 160.20%#			
50) Toluene-d8	10.185	98	80514	31.78	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 63.56%			
62) 4-Bromofluorobenzene	12.483	95	19151	22.83	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 45.66%			
Target Compounds						
						Qvalue
16) Acetone	3.979	43	43784	317.13	ug/l	94
18) Methyl Acetate	4.497	43	13515	37.43	ug/l	92
20) Methylene Chloride	4.728	84	74257	147.52	ug/l	95
25) 2-Butanone	7.008	43	5781	28.33	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
Data File : VY05576.D
Acq On : 06 Aug 2021 17:21
Operator : SY/MD
Sample : M3308-04
Misc : 1.19G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11930

Quant Time: Aug 07 06:20:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

