

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
 Data File : VY005577.D
 Acq On : 06 Aug 2021 17:44
 Operator : SY/MD
 Sample : M3289-09RE
 Misc : 1.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 37702RTRE

Quant Time: Aug 07 06:20:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	62615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	112372	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	47263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	12011	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	65527	103.38	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 206.76%#	
35) Dibromofluoromethane	7.722	113	44504	70.74	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 141.48%	
50) Toluene-d8	10.185	98	59721	22.63	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 45.26%#	
62) 4-Bromofluorobenzene	12.483	95	14135	16.18	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 32.36%	
Target Compounds						
						Qvalue
16) Acetone	3.954	43	5414	33.88	ug/l	96
18) Methyl Acetate	4.497	43	1157	2.77	ug/l	93
20) Methylene Chloride	4.729	84	31317	49.37	ug/l	90
25) 2-Butanone	7.003	43	1809	7.66	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080621\
Data File : VY005577.D
Acq On : 06 Aug 2021 17:44
Operator : SY/MD
Sample : M3289-09RE
Misc : 1.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
37702RTRE

Quant Time: Aug 07 06:20:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

